WSCF Laboratory
PO Box 650 S3-30
Richland, WA 99352

December 18, 2012

Scot Fitzgerald
CH2M-HILL PRC
PO Box 1600
Richland, WA 99352

Dear Scot Fitzgerald,

REVISED121232 - 698825 [Report ID: 121232]

Reference: (1) SOW, Mod 2, #36587, Release 3
(2) MSC-SD-CD-QAPP-017, current version, Waste Sampling & Characterization Facility Quality Assurance Program Plan

This letter contains the following information for sample delivery group WSCF121232

* Cover Sheet (Attachment 1)
* Narrative (Attachment 2)
* Analytical Results (Attachment 3)
* Sample Receipt Information (Attachment 4)

Very truly yours,

Electronically signed by Joseph Hale
For Lab Manager, Dan T. Smith
WSCF Analytical Lab
(509) 373-4804

Attachments 4
CC: w/Attachments

File/LB

ATTACHMENT 1

COVER SHEET

Consisting of 2 pages
Including cover page

WSCF SAF Number Cross Reference

Group # WSCF121232
Data Deliverable Date 11/05/12

SAF #	Sample ID	Sample #	Matrix	Sampled	Received
W13-010	B2M177	121232001	WATER	10/03/12	10/03/12
W13-010	B2M9W6	121232002	WATER	10/03/12	10/03/12
W13-010	B2M179	121232003	WATER	10/03/12	10/03/12
W13-010	B2M9W5	121232004	WATER	10/03/12	10/03/12
W13-010	B2M239	121232005	WATER	10/03/12	10/03/12
W13-010	B2M240	121232006	WATER	10/03/12	10/03/12
W13-010	B2M241	121232007	WATER	10/03/12	10/03/12
W13-010	B2M9W7	121232008	WATER	10/03/12	10/03/12
W13-010	B2M9W8	121232009	WATER	10/03/12	10/03/12
W13-010	B2M9W9	121232010	WATER	10/03/12	10/03/12
W13-010	B2M8K7	121232011	WATER	10/03/12	10/03/12
W13-010	B2M8K8	121232012	WATER	10/03/12	10/03/12

ATTACHMENT 2

NARRATIVE

Consisting of 8 pages
Including cover page

Attachment 2
Narrative Rev2
WSCF121232

Revision 2: This case narrative replaces the prior in its entirety. 1,4-Dioxane was added per SDR13-064 to samples B2M177 and B2M9W6.

Revision 1 P&D: Replaces prior case narrative in its entirety. The mercury result was for sample B2M9W6, the Kerosene results for samples B2M177 and B2M9W6 have been added to the report.

Introduction

Samples were received at the WSCF laboratory as referenced on the WSCF SAF Number Cross Reference table included in the final report. The samples were analyzed for the analytes indicated on the attached copy of the chain of custody (COC) form in accordance with the *Statement of Work (SOW), Master Contract 39818, Revision 3, "Laboratory Analytical Services to CHPRC Soil and Groundwater Remediation Project."*

The narrative (Attachment 2) will address sample characteristics, analyses requested and general information in performance of the analytical methods. A Data Summary Report (Attachment 3) includes analytical results, a comment report detailing method abnormalities, tentatively identified peaks if applicable, method references, and Laboratory QC information as applicable. Copies of the chain of custody and sample receipt documentation are included as Attachment 4.

It should be noted that the attached chain of custody was not stamped "ICED" by the WSCF Laboratory Sample Custodian during sample receiving. However, based on procedure LO-090-403 form "NOTICE OF IMPROPER SAMPLE SUBMITTAL" was not submitted and was not stamped "NOT ICED". No anomaly was noted during sample receipt.

- Sample Issue Resolution Form SDR13-012 regarding Hexachlorophene analysis by Method 8270 is attached to this report.
- Sample Issue Resolution Form SDR13-064 adding 1,4-Dioxane by 8270 is attached to this report.

The following generic data qualifiers (i.e., B, C, D, J and U) may be applicable to this report, as appropriate.

- **B** – Sample results with a concentration greater than the MDL but less than the PQL are B flagged (applies to inorganic and wet chemical analyses), as appropriate.
- **C** – Analyte was detected in the blank and was evaluated. Affected sample results in the batch were C flagged (applies to inorganic and wet chemical analyses).
- **D** – Sample results are D flagged if dilution(s) were required, as appropriate.
- **J** – Sample results with a concentration greater than the MDL but less than the PQL are J flagged (applies to organic analyses), as appropriate.

Attachment 2
Narrative Rev2
WSCF121232

- **B (organic analyses)** – Analyte was detected in the blank and was evaluated. Affected sample results in the batch were B flagged.

- **U** – Analyzed for but not detected above limiting criteria. Relative Percent Difference (RPD) values associated with an analyte qualified with a "U" are not applicable.

Analytical Methodology for Requested Analyses

Refer to *WSCF Method References Report* for a complete listing of approved analytical methods.

Inorganic Comments

Chemical Oxygen Demand – Hold time requirement for this analysis was met. A Duplicate, Matrix Spike, Matrix Spike Duplicate and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

Cyanide – The hold time requirement for this analysis was met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

ICP-AES Metals – The hold time requirements for this analysis were met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

ICP-MS Metals – The hold time requirements for this analysis were met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- Batch QC 208454
  - Vanadium, Strontium, Barium, Molybdenum and Lead were detected in the Blank and evaluated.
- Batch QC 208257
  - Vanadium, Strontium, Copper and Aluminum were detected in the Blank and evaluated.
- Batch QC 208289
  - Vanadium was detected in the Blank and evaluated.

Attachment 2
Narrative Rev2
WSCF121232

- All other applicable QC controls are within the established limits.

Total Alkalinity – The hold time requirement for this analysis was met. A Duplicate and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

Total Organic Carbon – The hold time requirement for this analysis was met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

Total Organic Halides – The hold time requirement for this analysis was met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

Organic Comments

PCB – The hold time requirements for this analysis were met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

Semi-VOA – The hold time requirements for this analysis were met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- 4-Nitrophenol and Pentachlorophenol did not meet the MS / MSD RPD acceptance limits. Sample results for this analyte were not flagged. The quality control report was flagged for RPD failure.
- The Blank, MS, MSD and samples B2M177 (121232001) and B2M9W6 (121232002) did not meet the acceptance limits for surrogate Phenol-d5. Sample results were not flagged. The quality control report was flagged for surrogate recovery failure.
- All other applicable QC controls are within the established limits.

TPHD-WA – The hold time requirements for this analysis were met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

Attachment 2
Narrative Rev2
WSCF121232

TPHG-WA – The hold time requirement for this analysis was met. A Duplicate, Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

VOA – The hold time requirement for this analysis was met. A Matrix Spike, Matrix Spike Duplicate, Blank and Laboratory Control Sample were analyzed with this delivery group. Analytical Note(s):

- All applicable QC controls are within the established limits.

We certify that this data package is in compliance with the SOW, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this data package has been authorized by the Analytical Laboratory Manager (or designee) and the Client Services representative as verified by electronic signatures shown on the WSCF ANALYTICAL RESULTS REPORT.

Attachment 2
Narrative Rev2
WSCF121232

SAMPLE ISSUE RESOLUTION

SIR NUM	SDR13-012
REV NUM	0
DATE INITIATED	10/15/2012

SAMPLE EVENT INFORMATION

SAF NUM(S)	W13-010
OPERABLE UNIT(S)	
PROJECT(S)	RCRA13
SAMPLE EVENT TITLE(S)	RCRA13
LABORATORY	Waste Sampling & Characterization

SAMPLING INFORMATION

NUMBER OF SAMPLES	18
SAMPLE NUMBERS	B2M0Y3, B2M0Y9, B2M105, B2M111, B2M117, B2M123, B2M129, B2M135, B2M141, B2M147, B2M153, B2M159, B2M165, B2M171, B2M177, B2M194, B2M1B0, B2M9W6
SAMPLE MATRIX	WATER
COLLECTION DATE	10/3/2012 - 10/10/2012
SDG NUM	WSCF121241, WSCF121242, WSCF121284, WSCF121226, WSCF121275, WSCF121230, WSCF121223, WSCF121274, WSCF121232, WSCF121239

ISSUE BACKGROUND

CLASS	Laboratory Issue
TYPE	Cancellation of Analyses
DESCRIPTION	WSCF is not currently calibrated for Hexachlorophene by 8270. The compound is polar thus it sticks to the columns. Therefore, it is not able to be routinely measure in the extract. In order to be able to report the compound it would need to be derivatized. WSCF is currently not setup to derivatize Hexachlorophene.

DISPOSITION

DESCRIPTION	PROPOSED DISPOSITION: Report the data without Hexachlorophene and note issue in narrative.
JUSTIFICATION	ACCEPTED DISPOSITION: Accept proposed resolution and request that WSCF send a weekly update of any further SDGs that are affected to be added to this SIR. SUBMITTED BY: Heather Medley/WSCF DATE: 10/15/12 ACCEPTED BY: Karen Waters-Husted/CHPRC DATE: 10/16/12

A-6003-640 (03/30) PAGE 1 of 1

Attachment 2
Narrative Rev2
WSCF121232

Problem and Discrepancy Report

WSCF

SDG WSCF121232

11/06/2012

---

1. The data package has the following issues:

a) TPHKEROSENE for sample number B2M177 and B2M9W6was not reported in the electronic or hardcopy data packages.
b) Mercury (200.8_HG-ICMS) was not reported for sample 2BM9W6.

Resolution: *Provide appropriate correction*

Lab Response: **results have been added.**

Attachment 2
Narrative Rev2
WSCF121232

SAMPLE ISSUE RESOLUTION

SIR NUM	SDR13-064
REV NUM	0
DATE INITIATED	12/12/2012

SAMPLE EVENT INFORMATION

SAF NUM(S)	W13-011, X13-012, W13-012, I13-007, W13-002, W13-010
OPERABLE UNIT(S)	NONE, 100-NR-2
PROJECT(S)	SURV13, CERC13, RCRA13
SAMPLE EVENT TITLE(S)	SURV13, CERC13, RCRA13
LABORATORY	Waste Sampling & Characterization

SAMPLING INFORMATION

NUMBER OF SAMPLES	63
SAMPLE NUMBERS	B2M0Y3, B2M0Y9, B2M105, B2M111, B2M117, B2M123, B2M129, B2M135, B2M141, B2M147, B2M153, B2M159, B2M165, B2M171, B2M177, B2M183, B2M194, B2M1B0, B2M9W6, B2MN27, B2MN32, B2MN37, B2MN42, B2MN47, B2MN52, B2MN57, B2MN64, B2MN71, B2MN75, B2MN81, B2MN82, B2MN89, B2MN94, B2MN99, B2MNB5, B2MNC0, B2MXD6, B2MXL9, B2MXM7, B2MXM8, B2MXN9, B2N3D3, B2N3D5, B2N905, B2N906, B2N910, B2N914, B2N915, B2N919, B2N923, B2N926, B2N931, B2N935, B2N938, B2N941, B2N945, B2N946, B2N950, B2N953, B2N957, B2N958, B2N962, B2N965
SAMPLE MATRIX	WATER
COLLECTION DATE	10/3/2012 - 12/6/2012
SDG NUM	WSCF121241, WSCF121242, WSCF121404, WSCF121448, WSCF121284, WSCF121439, WSCF121226, WSCF121555, WSCF121456, WSCF121275, WSCF121230, WSCF121223, WSCF121303, WSCF121402, WSCF121401, WSCF121398, WSCF121274, WSCF121232, WSCF121450, WSCF121239

ISSUE BACKGROUND

CLASS	Sample Management Issues
TYPE	Addition of Analyses
DESCRIPTION	Missed adding Appendix IX constituent 1,4-Dioxane to the new service list.
DISPOSITION DESCRIPTION	Proposed disposition: Request that WSCF report the missing data for 1,4-Dioxane for the listed samples. SMR will add necessary information to the affected data packages.
JUSTIFICATION	Accepted disposition: WSCF understands SMR missed having 1,4-dioxane reported for the Appendix IX 8270 service list. The data is available. WSCF has added 1,4-dioxane to the yellow highlighted samples above in addition to B2MXN8 (121450), and B2MN63 (121456). The samples not highlighted have not been received by WSCF as of 12/17/12. When they are received 1,4-dioxane will be added.

Submitted by: Karen Waters-Husted/CHPRC DATE: 12/12/12
Accepted by: Heather Medley/WSCF DATE: 12/17/12

A-6003-640 (03/30) PAGE 1 of 1

ATTACHMENT 3

ANALYTICAL RESULTS

Consisting of 123 pages
Including cover page

WSCF ANALYTICAL RESULTS REPORT

For

CH2M Hill Plateau Remediation

PO Box 1600
Richland, WA 99352

Attention: Scot Fitzgerald

Contract # MOA-FH-CHPRC-2008
Group # WSCF121232
Report Date December 18, 2012

Analytical: Electronically signed by Joseph Hale

Client Services: Electronically signed by Heather Medley

Solid samples results that have a 'Percent Solid' test are reported on a "dry weight basis", except results of TCLP, Percent Solid, and Total Activity. If no `Percent Solid' test is reported then the results are reported on an "as received" basis.

Batch QC List

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

QC Batch	Analytical Batch	S#	Type	Sample #	Client Sample#	Original	Test
208198	208205	5	BLANK	82193	BLANK		ICP-6010 - All possible metals
208198	208205	7	LCS	82195	LCS		ICP-6010 - All possible metals
208198	208205	9	MS	82196	B2M131(121223007MS)	121223007	ICP-6010 - All possible metals
208198	208205	10	MSD	82197	B2M131(121223007MSD)	121223007	ICP-6010 - All possible metals
208198	208205	22	SAMPLE	121232003	B2M179		ICP-6010 - All possible metals
208198	208205	23	SAMPLE	121232004	B2M9W5		ICP-6010 - All possible metals
208198	208205	24	SAMPLE	121232001	B2M177		ICP-6010 - All possible metals
208198	208205	25	SAMPLE	121232002	B2M9W6		ICP-6010 - All possible metals
208247	208248	5	BLANK	82359	BLANK		Chemical Oxygen Demand
208247	208248	6	LCS	82360	LCS		Chemical Oxygen Demand
208247	208248	9	MS	82363	B2M1M4(121215007MS)	121215007	Chemical Oxygen Demand
208247	208248	10	MSD	82364	B2M1M4(121215007MSD	121215007	Chemical Oxygen Demand
208247	208248	16	SAMPLE	121232001	B2M177		Chemical Oxygen Demand
208247	208248	17	SAMPLE	121232002	B2M9W6		Chemical Oxygen Demand
208257	208292	4	BLANK	82385	BLANK		ICP-2008 MS All possible metal
208257	208292	5	LCS	82386	LCS		ICP-2008 MS All possible metal
208257	208292	7	MS	82387	B2M8H4(121223010MS)	121223010	ICP-2008 MS All possible metal
208257	208292	8	MSD	82388	B2M8H4(121223010MSD	121223010	ICP-2008 MS All possible metal
208257	208292	28	SAMPLE	121232003	B2M179		ICP-2008 MS All possible metal
208289	208303	4	BLANK	82511	BLANK		ICP-2008 MS All possible metal
208289	208303	5	LCS	82512	LCS		ICP-2008 MS All possible metal
208289	208303	6	SAMPLE	121232004	B2M9W5		ICP-2008 MS All possible metal
208289	208303	7	MS	82513	B2M9W5(121232004MS)	121232004	ICP-2008 MS All possible metal

REVISED121232 -

Batch QC List

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

QC Batch	Analytical Batch	S#	Type	Sample #	Client Sample#	Original	Test
208289	208303	8	MSD	82514	B2M9W5(121232004MSD	121232004	ICP-2008 MS All possible metal
208289	208303	9	SAMPLE	121232011	B2M8K7		ICP-2008 MS All possible metal
208289	208303	10	SAMPLE	121232012	B2M8K8		ICP-2008 MS All possible metal
208454	208492	4	BLANK	82681	BLANK		ICP-2008 MS All possible metal
208454	208492	5	LCS	82682	LCS		ICP-2008 MS All possible metal
208454	208492	7	MS	82683	B2MC52(121283007MS)	121283007	ICP-2008 MS All possible metal
208454	208492	8	MSD	82684	B2MC52(121283007MSD	121283007	ICP-2008 MS All possible metal
208454	208492	23	SAMPLE	121232001	B2M177		ICP-2008 MS All possible metal
208454	208492	24	SAMPLE	121232002	B2M9W6		ICP-2008 MS All possible metal
209136	209139	1	BLANK	83612	BLANK		Total Organic Halides
209136	209139	2	LCS	83613	LCS		Total Organic Halides
209136	209139	20	MS	83617	B2M177(121232001MS)	121232001	Total Organic Halides
209136	209139	21	MSD	83618	B2M177(121232001MSD)	121232001	Total Organic Halides
209136	209139	22	SAMPLE	121232001	B2M177		Total Organic Halides
209136	209139	23	SAMPLE	121232002	B2M9W6		Total Organic Halides
209136	209139	24	SAMPLE	121232005	B2M239		Total Organic Halides
209136	209139	25	SAMPLE	121232006	B2M240		Total Organic Halides
209136	209139	26	SAMPLE	121232007	B2M241		Total Organic Halides
209136	209139	27	SAMPLE	121232008	B2M9W7		Total Organic Halides
209136	209139	28	SAMPLE	121232009	B2M9W8		Total Organic Halides
209136	209139	29	SAMPLE	121232010	B2M9W9		Total Organic Halides

Batch QC List

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

QC Batch	Analytical Batch	S#	Type	Sample #	Client Sample#	Original	Test
208245	208258	1	BLANK	82350	BLANK		Extractable Diesel and Petroleum
208245	208258	2	LCS	82351	LCS		Extractable Diesel and Petroleum
208245	208258	3	MS	82352	B2M129(121223013MS)	121223013	Extractable Diesel and Petroleum
208245	208258	4	MSD	82353	B2M129(121223013MSD)	121223013	Extractable Diesel and Petroleum
208245	208258	9	SAMPLE	121232002	B2M9W6		Extractable Diesel and Petroleum
208245	208258	11	SAMPLE	121232001	B2M177		Extractable Diesel and Petroleum
208488	208850	1	BLANK	82826	BLANK		SW-846 8270D Semivolatiles
208488	208850	2	LCS	82827	LCS		SW-846 8270D Semivolatiles
208488	208850	3	MS	82828	B2M0Y3(121239001MS)	121239001	SW-846 8270D Semivolatiles
208488	208850	4	MSD	82829	B2M0Y3(121239001MSD	121239001	SW-846 8270D Semivolatiles
208488	208850	9	SAMPLE	121232002	B2M9W6		SW-846 8270D Semivolatiles
208488	208850	11	SAMPLE	121232001	B2M177		SW-846 8270D Semivolatiles
208650	208887	1	BLANK	82953	BLANK		PCBs by EPA SW-846 Method 8082
208650	208887	2	LCS	82954	LCS		PCBs by EPA SW-846 Method 8082
208650	208887	3	MS	82955	B2M129(121223013MS)	121223013	PCBs by EPA SW-846 Method 8082
208650	208887	4	MSD	82956	B2M129(121223013MSD)	121223013	PCBs by EPA SW-846 Method 8082
208650	208887	7	SAMPLE	121232002	B2M9W6		PCBs by EPA SW-846 Method 8082
208650	208887	9	SAMPLE	121232001	B2M177		PCBs by EPA SW-846 Method 8082

Batch QC List

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

QC Batch	Analytical Batch	S#	Type	Sample #	Client Sample#	Original	Test
208456	208457	1	BLANK	82689	BLANK		SW-846 8260B Volatiles
208456	208457	2	LCS	82690	LCS		SW-846 8260B Volatiles
208456	208457	3	MS	82691	B2M129(121223013MS)	121223013	SW-846 8260B Volatiles
208456	208457	4	MSD	82692	B2M129(121223013MSD)	121223013	SW-846 8260B Volatiles
208456	208457	9	SAMPLE	121232002	B2M9W6		SW-846 8260B Volatiles
208456	208457	11	SAMPLE	121232001	B2M177		SW-846 8260B Volatiles
208756	208757	1	BLANK	82983	BLANK		Gasoline Range (W)
208756	208757	2	LCS	82984	LCS		Gasoline Range (W)
208756	208757	3	MS	82985	B2M129(121223013MS)	121223013	Gasoline Range (W)
208756	208757	4	MSD	82986	B2M129(121223013MSD)	121223013	Gasoline Range (W)
208756	208757	5	DUP	82987	B2M129(121223013DUP)	121223013	Gasoline Range (W)
208756	208757	15	SAMPLE	121232002	B2M9W6		Gasoline Range (W)
208756	208757	16	SAMPLE	121232001	B2M177		Gasoline Range (W)

Batch QC List

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

QC Batch	Analytical Batch	S#	Type	Sample #	Client Sample#	Original	Test
208268	208268	2	BLANK	82439	BLANK		Total Organic Carbon
208268	208268	3	LCS	82440	LCS		Total Organic Carbon
208268	208268	17	MS	82444	B2M213(121230004MS)	121230004	Total Organic Carbon
208268	208268	18	MSD	82445	B2M213(121230004MSD)	121230004	Total Organic Carbon
208268	208268	21	SAMPLE	121232001	B2M177		Total Organic Carbon
208268	208268	22	SAMPLE	121232002	B2M9W6		Total Organic Carbon
208268	208268	23	SAMPLE	121232005	B2M239		Total Organic Carbon
208268	208268	24	SAMPLE	121232006	B2M240		Total Organic Carbon
208268	208268	25	SAMPLE	121232007	B2M241		Total Organic Carbon
208268	208268	26	SAMPLE	121232008	B2M9W7		Total Organic Carbon
208268	208268	27	SAMPLE	121232009	B2M9W8		Total Organic Carbon
208268	208268	28	SAMPLE	121232010	B2M9W9		Total Organic Carbon
208476	208482	1	BLANK	82749	BLANK		Cyanide (W) by Midi/Spectrophotometer
208476	208482	4	LCS	82752	LCS		Cyanide (W) by Midi/Spectrophotometer
208476	208482	5	MS	82753	B2LDT5(121203010MS)	121203010	Cyanide (W) by Midi/Spectrophotometer
208476	208482	6	MSD	82754	B2LDT5(121203010MSD)	121203010	Cyanide (W) by Midi/Spectrophotometer
208476	208482	17	SAMPLE	121232001	B2M177		Cyanide (W) by Midi/Spectrophotometer
208476	208482	18	SAMPLE	121232002	B2M9W6		Cyanide (W) by Midi/Spectrophotometer
208754	208754	1	LCS	82975	LCS		Total Alkalinity as mg/L CaCO3 (Water)
208754	208754	4	SAMPLE	121232001	B2M177		Total Alkalinity as mg/L CaCO3 (Water)
208754	208754	5	SAMPLE	121232002	B2M9W6		Total Alkalinity as mg/L CaCO3 (Water)
208754	208754	9	DUP	82976	B2M0Y9(121239002DUP)	121239002	Total Alkalinity as mg/L CaCO3 (Water)
208754	208754	13	LCS	82977	LCS		Total Alkalinity as mg/L CaCO3 (Water)

REVISED121232 -

Batch QC List

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

QC Batch	Analytical Batch	S#	Type	Sample #	Client Sample#	Original	Test
208754	208754	24	LCS	82978	LCS		Total Alkalinity as mg/L CaCO3 (Water)

Method Reference

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

The results provided in this report were generated using the following WSCF Laboratory procedures. For your convenience, this table provides a listing of the regulatory, industry methods or HEIS methods that are referenced by each of these WSCF procedures. Please note that the most recent version of the regulatory or industry method is listed here even though WSCF procedure may reference an older version of the method. Also, a reference to a regulatory or industry method here does not necessarily indicate a verbatim implementation of that method.

LA-505-411	Elemental Analysis by ICP Atomic Emission Spectroscopy (ICP AES)		
	EPA SW-846	6010C	Inductively Coupled Plasma-Atomic Emmision Spectrometry
	HEIS	6010_METALS_ICP	Inductively Coupled Plasma-Atomic Emmision Spectrometry
LA-505-412	Determination of Trace Elements in Waters & Wastes by ICP Mass Spectrometry		
	EPA-600/R-94-111	200.8	Determination of Trace Elements in Waters and Waste by Inductively Coupled Plasma
	HEIS	200.8_METALS_ICPMS	Determination of Trace Elements in Waters and Waste by Inductively Coupled Plasma, Mass Spec.
LA-523-444	Total Organic Halides Based on SW-846 Method 9020B		
	EPA SW-846	9020B	Total Organic Halides (TOX)
	HEIS	9020_TOX	Total Organic Halides (TOX)
LA-523-470	Chemical Oxygen Demand		
	EPA-600/4-79-020	410.4	Chemical Oxygen Demand
	HEIS	410.4_COD	Chemical Oxygen Demand

Note: A complete list of WSCF analytical procedures and reference regulatory or industry methods is available online at http://www7.rl.gov/rapidweb/AS-DOL/index.cfm

Method Reference

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

The results provided in this report were generated using the following WSCF Laboratory procedures. For your convenience, this table provides a listing of the regulatory, industry methods or HEIS methods that are referenced by each of these WSCF procedures. Please note that the most recent version of the regulatory or industry method is listed here even though WSCF procedure may reference an older version of the method. Also, a reference to a regulatory or industry method here does not necessarily indicate a verbatim implementation of that method.

LA-523-427	Polychlorinated Biphenyls (PCB'S) by Gas Chromatography		
	EPA SW-846	3510C	Separatory Funnel Liquid-Liquid Extraction
	EPA SW-846	3545	Pressurized Fluid Extraction (PFE)
	EPA SW-846	3665A	Sulfuric Acid/Permanganate Cleanup
	EPA SW-846	8000B	Determinative Chromatographic Separations
	EPA SW-846	8082	Polychlorinated Biphenyls (PCBs) by Gas Chromatography
	HEIS	8082_PCB_GC	Polychlorinated Biphenyls (PCBs) by Gas Chromatography
LA-523-456	Semivolatile Sample Analysis by SW-846 Method 8270D		
	EPA SW-846	8000B	Determinative Chromagraphic Separations
	EPA SW-846	3510C	Separatory Funnel Liquid-Liquid Extraction
	EPA SW-846	8270D	Semivolatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS)
	EPA SW-846	3545	Pressurized Fluid Extraction (PFE)
	HEIS	8270_SVOA_GCMS	Semivolatile Organic Compounds by Gas Chromatography/Mass Spectrometry(GC/MS)
LA-523-493	NWTPH-Dx, Extractible Diesel and Petroleum Productions Analysis in Soil and Water		
	WDOE	WDOE	Total Petroleum Hydrocarbons in Diesel
	HEIS	WTPH_DIESEL	TPH Diesel

Note: A complete list of WSCF analytical procedures and reference regulatory or industry methods is available online at http://www7.rl.gov/rapidweb/AS-DOL/index.cfm

Method Reference

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

The results provided in this report were generated using the following WSCF Laboratory procedures. For your convenience, this table provides a listing of the regulatory, industry methods or HEIS methods that are referenced by each of these WSCF procedures. Please note that the most recent version of the regulatory or industry method is listed here even though WSCF procedure may reference an older version of the method. Also, a reference to a regulatory or industry method here does not necessarily indicate a verbatim implementation of that method.

LA-523-443	Gas Chromatography Analysis of Gasoline Range Total Petroleum Hydrocarbons in Water		
	EPA	NWTPH-G	"Analytical Methods for Petroleum Hydrocarbons, June 1997, NWTPH-G, Volatile Petroleum Products Method for Soil and Water
	HEIS	WTPH_GASOLINE	Total Petroleum Hydrocarbons, Gasoline
LA-523-455	Volatile Sample Analysis by SW-846 Method 8260B		
	EPA SW-846	8000B	Determinative Chromographic Separations
	EPA SW-846	8260B	Volatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS)
	HEIS	8260_VOA_GCMS	Volatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS)

Note: A complete list of WSCF analytical procedures and reference regulatory or industry methods is available online at http://www7.rl.gov/rapidweb/AS-DOL/index.cfm

Method Reference

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

The results provided in this report were generated using the following WSCF Laboratory procedures. For your convenience, this table provides a listing of the regulatory, industry methods or HEIS methods that are referenced by each of these WSCF procedures. Please note that the most recent version of the regulatory or industry method is listed here even though WSCF procedure may reference an older version of the method. Also, a reference to a regulatory or industry method here does not necessarily indicate a verbatim implementation of that method.

LA-531-411	Alkalinity		
	SM	2320	Alkalinity
	HEIS	2320_ALKALINITY	Alkalinity
LA-344-406	Total Organic Carbon (TOC) Based on SW-846		
	EPA SW-846	9060	Total Organic Carbon
	HEIS	9060_TOC	Total Organic Carbon
LA-695-402	Determination of Cyanide by Mididistillation and		
	EPA	SW-846 Method 9014/9010	Determination of Cyanide by Midi-Distillation and Spectrophotometric Analysis
	SM	4500 CNE	Cyanide, Total
	HEIS	4500E_CN	Cyanide, Total

Note: A complete list of WSCF analytical procedures and reference regulatory or industry methods is available online at http://www7.rl.gov/rapidweb/AS-DOL/index.cfm

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
ICPAES Prep (W)										**10/04/12**
ICP-6010 - All possible metals										
Iron	7439-89-6	LA-505-411	B	67.2		ug/L	1	19	95	10/05/12
Magnesium	7439-95-4	LA-505-411		26400		ug/L	1	4.0	20	10/05/12
Manganese	7439-96-5	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Nickel	7440-02-0	LA-505-411	B	12.0		ug/L	1	4.0	20	10/05/12
Potassium	7440-09-7	LA-505-411		9740		ug/L	1	76	380	10/05/12
Silver	7440-22-4	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Sodium	7440-23-5	LA-505-411		33100		ug/L	1	10	50	10/05/12
Antimony	7440-36-0	LA-505-411	U	<36		ug/L	1	36	180	10/05/12
Barium	7440-39-3	LA-505-411		99.3		ug/L	1	4.0	20	10/05/12
Cadmium	7440-43-9	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Chromium	7440-47-3	LA-505-411	B	13.3		ug/L	1	5.0	25	10/05/12
Cobalt	7440-48-4	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Copper	7440-50-8	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Vanadium	7440-62-2	LA-505-411	B	10.8		ug/L	1	5.0	25	10/05/12
Zinc	7440-66-6	LA-505-411	U	<5.0		ug/L	1	5.0	25	10/05/12
Calcium	7440-70-2	LA-505-411		89500		ug/L	1	49	240	10/05/12
Strontium	7440-24-6	LA-505-411		441		ug/L	1	9.0	45	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Titanium	7440-32-6	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Beryllium	7440-41-7	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
ICPMS Prep (W)										**10/12/12**
ICP-2008 MS All possible metal										
Aluminum	7429-90-5	LA-505-412	D	607		ug/L	2	10	100	10/15/12
Manganese	7439-96-5	LA-505-412	D	2.20		ug/L	2	0.20	2.0	10/15/12
Nickel	7440-02-0	LA-505-412	D	9.46		ug/L	2	0.20	2.0	10/15/12
Silver	7440-22-4	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/15/12
Antimony	7440-36-0	LA-505-412	UD	<0.60		ug/L	2	0.60	6.0	10/15/12
Barium	7440-39-3	LA-505-412	D	109		ug/L	2	0.40	4.0	10/15/12
Beryllium	7440-41-7	LA-505-412	UD	<0.20		ug/L	2	0.20	2.0	10/15/12
Cadmium	7440-43-9	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/15/12
Chromium	7440-47-3	LA-505-412	D	14.5		ug/L	2	0.20	2.0	10/15/12
Cobalt	7440-48-4	LA-505-412	BD	0.280		ug/L	2	0.10	0.50	10/15/12
Copper	7440-50-8	LA-505-412	BD	0.792		ug/L	2	0.20	2.0	10/15/12
Vanadium	7440-62-2	LA-505-412	DC	17.6		ug/L	2	0.40	4.0	10/15/12
Zinc	7440-66-6	LA-505-412	BD	10.3		ug/L	2	2.0	20	10/15/12
Lead	7439-92-1	LA-505-412	BDC	0.568		ug/L	2	0.10	1.0	10/15/12
Molybdenum	7439-98-7	LA-505-412	D	4.49		ug/L	2	0.10	1.0	10/15/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Strontium	7440-24-6	LA-505-412	D	525		ug/L	2	0.20	2.0	10/15/12
Thallium	7440-28-0	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/15/12
Tin	7440-31-5	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/15/12
Arsenic	7440-38-2	LA-505-412	D	5.45		ug/L	2	0.40	4.0	10/15/12
Selenium	7782-49-2	LA-505-412	BD	7.63		ug/L	2	2.0	20	10/15/12
Preparation for COD (W)										**10/04/12**
Chemical Oxygen Demand										
Chemical Oxygen Demand	COD	LA-523-470	U	<10		mg/L	1	10	50	10/04/12
Preparation for TOX (W)										**10/08/12**
Total Organic Halides										
Total Organic Halides	59473-04-0	LA-523-444	U	<5.0		ug/L	1	5.0	15	10/08/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
ICPAES Prep (W)										**10/04/12**
ICP-6010 - All possible metals										
Iron	7439-89-6	LA-505-411	B	49.2		ug/L	1	19	95	10/05/12
Magnesium	7439-95-4	LA-505-411		20100		ug/L	1	4.0	20	10/05/12
Manganese	7439-96-5	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Nickel	7440-02-0	LA-505-411	B	7.40		ug/L	1	4.0	20	10/05/12
Potassium	7440-09-7	LA-505-411		8620		ug/L	1	76	380	10/05/12
Silver	7440-22-4	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Sodium	7440-23-5	LA-505-411		28800		ug/L	1	10	50	10/05/12
Antimony	7440-36-0	LA-505-411	U	<36		ug/L	1	36	180	10/05/12
Barium	7440-39-3	LA-505-411		57.8		ug/L	1	4.0	20	10/05/12
Cadmium	7440-43-9	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Chromium	7440-47-3	LA-505-411	B	9.60		ug/L	1	5.0	25	10/05/12
Cobalt	7440-48-4	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Copper	7440-50-8	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Vanadium	7440-62-2	LA-505-411	B	17.7		ug/L	1	5.0	25	10/05/12
Zinc	7440-66-6	LA-505-411	U	<5.0		ug/L	1	5.0	25	10/05/12
Calcium	7440-70-2	LA-505-411		66600		ug/L	1	49	240	10/05/12
Strontium	7440-24-6	LA-505-411		331		ug/L	1	9.0	45	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Titanium	7440-32-6	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Beryllium	7440-41-7	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
ICPMS Prep (W)										**10/12/12**
ICP-2008 MS All possible metal										
Aluminum	7429-90-5	LA-505-412	BD	46.7		ug/L	2	10	100	10/15/12
Manganese	7439-96-5	LA-505-412	BD	0.962		ug/L	2	0.20	2.0	10/15/12
Nickel	7440-02-0	LA-505-412	D	4.73		ug/L	2	0.20	2.0	10/15/12
Silver	7440-22-4	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/15/12
Antimony	7440-36-0	LA-505-412	UD	<0.60		ug/L	2	0.60	6.0	10/15/12
Barium	7440-39-3	LA-505-412	D	60.9		ug/L	2	0.40	4.0	10/15/12
Beryllium	7440-41-7	LA-505-412	UD	<0.20		ug/L	2	0.20	2.0	10/15/12
Cadmium	7440-43-9	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/15/12
Chromium	7440-47-3	LA-505-412	D	9.84		ug/L	2	0.20	2.0	10/15/12
Cobalt	7440-48-4	LA-505-412	BD	0.154		ug/L	2	0.10	0.50	10/15/12
Copper	7440-50-8	LA-505-412	BD	0.536		ug/L	2	0.20	2.0	10/15/12
Vanadium	7440-62-2	LA-505-412	D	19.0		ug/L	2	0.40	4.0	10/15/12
Zinc	7440-66-6	LA-505-412	UD	<2.0		ug/L	2	2.0	20	10/15/12
Lead	7439-92-1	LA-505-412	BDC	0.216		ug/L	2	0.10	1.0	10/15/12
Mercury	7439-97-6	LA-505-412	U	<0.10		ug/L	2	0.10	0.40	10/15/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Molybdenum	7439-98-7	LA-505-412	D	6.79		ug/L	2	0.10	1.0	10/15/12
Strontium	7440-24-6	LA-505-412	D	367		ug/L	2	0.20	2.0	10/15/12
Thallium	7440-28-0	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/15/12
Tin	7440-31-5	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/15/12
Arsenic	7440-38-2	LA-505-412	D	5.31		ug/L	2	0.40	4.0	10/15/12
Selenium	7782-49-2	LA-505-412	BD	6.84		ug/L	2	2.0	20	10/15/12
Preparation for COD (W)										**10/04/12**
Chemical Oxygen Demand										
Chemical Oxygen Demand	COD	LA-523-470	U	<10		mg/L	1	10	50	10/04/12
Preparation for TOX (W)										**10/08/12**
Total Organic Halides										
Total Organic Halides	59473-04-0	LA-523-444	U	<5.0		ug/L	1	5.0	15	10/08/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232003
SAF# W13-010
Sample ID B2M179

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
ICPAES Prep (W)										**10/04/12**
ICP-6010 - All possible metals										
Iron	7439-89-6	LA-505-411	B	46.8		ug/L	1	19	95	10/05/12
Magnesium	7439-95-4	LA-505-411		27000		ug/L	1	4.0	20	10/05/12
Manganese	7439-96-5	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Nickel	7440-02-0	LA-505-411	B	8.00		ug/L	1	4.0	20	10/05/12
Potassium	7440-09-7	LA-505-411		9970		ug/L	1	76	380	10/05/12
Silver	7440-22-4	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Sodium	7440-23-5	LA-505-411		33900		ug/L	1	10	50	10/05/12
Antimony	7440-36-0	LA-505-411	U	<36		ug/L	1	36	180	10/05/12
Barium	7440-39-3	LA-505-411		102		ug/L	1	4.0	20	10/05/12
Cadmium	7440-43-9	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Chromium	7440-47-3	LA-505-411	B	5.90		ug/L	1	5.0	25	10/05/12
Cobalt	7440-48-4	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Copper	7440-50-8	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Vanadium	7440-62-2	LA-505-411	B	14.5		ug/L	1	5.0	25	10/05/12
Zinc	7440-66-6	LA-505-411	U	<5.0		ug/L	1	5.0	25	10/05/12
Calcium	7440-70-2	LA-505-411		90600		ug/L	1	49	240	10/05/12
Strontium	7440-24-6	LA-505-411		443		ug/L	1	9.0	45	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232003
SAF# W13-010
Sample ID B2M179

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Titanium	7440-32-6	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Beryllium	7440-41-7	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
ICPMS Prep (W)										**10/09/12**
ICP-2008 MS All possible metal										
Aluminum	7429-90-5	LA-505-412	UD	<10		ug/L	2	10	100	10/09/12
Manganese	7439-96-5	LA-505-412	BD	1.75		ug/L	2	0.20	2.0	10/09/12
Nickel	7440-02-0	LA-505-412	D	7.40		ug/L	2	0.20	2.0	10/09/12
Silver	7440-22-4	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/09/12
Antimony	7440-36-0	LA-505-412	UD	<0.60		ug/L	2	0.60	6.0	10/09/12
Barium	7440-39-3	LA-505-412	D	99.1		ug/L	2	0.40	4.0	10/09/12
Beryllium	7440-41-7	LA-505-412	UD	<0.20		ug/L	2	0.20	2.0	10/09/12
Cadmium	7440-43-9	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/09/12
Chromium	7440-47-3	LA-505-412	D	6.19		ug/L	2	0.20	2.0	10/09/12
Cobalt	7440-48-4	LA-505-412	BD	0.298		ug/L	2	0.10	0.50	10/09/12
Copper	7440-50-8	LA-505-412	BDC	0.256		ug/L	2	0.20	2.0	10/09/12
Vanadium	7440-62-2	LA-505-412	D	13.9		ug/L	2	0.40	4.0	10/09/12
Zinc	7440-66-6	LA-505-412	UD	<2.0		ug/L	2	2.0	20	10/09/12
Lead	7439-92-1	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/09/12
Molybdenum	7439-98-7	LA-505-412	D	4.14		ug/L	2	0.10	1.0	10/09/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232003
SAF# W13-010
Sample ID B2M179

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Strontium	7440-24-6	LA-505-412	D	461		ug/L	2	0.20	2.0	10/09/12
Thallium	7440-28-0	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/09/12
Tin	7440-31-5	LA-505-412	UD	<0.10		ug/L	2	0.10	1.0	10/09/12
Arsenic	7440-38-2	LA-505-412	D	4.46		ug/L	2	0.40	4.0	10/09/12
Selenium	7782-49-2	LA-505-412	BD	7.01		ug/L	2	2.0	20	10/09/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232004
SAF# W13-010
Sample ID B2M9W5

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
ICPAES Prep (W)										**10/04/12**
ICP-6010 - All possible metals										
Iron	7439-89-6	LA-505-411	B	32.1		ug/L	1	19	95	10/05/12
Magnesium	7439-95-4	LA-505-411		20500		ug/L	1	4.0	20	10/05/12
Manganese	7439-96-5	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Nickel	7440-02-0	LA-505-411	B	6.30		ug/L	1	4.0	20	10/05/12
Potassium	7440-09-7	LA-505-411		8830		ug/L	1	76	380	10/05/12
Silver	7440-22-4	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Sodium	7440-23-5	LA-505-411		29600		ug/L	1	10	50	10/05/12
Antimony	7440-36-0	LA-505-411	U	<36		ug/L	1	36	180	10/05/12
Barium	7440-39-3	LA-505-411		59.1		ug/L	1	4.0	20	10/05/12
Cadmium	7440-43-9	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Chromium	7440-47-3	LA-505-411	B	5.30		ug/L	1	5.0	25	10/05/12
Cobalt	7440-48-4	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Copper	7440-50-8	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Vanadium	7440-62-2	LA-505-411	B	16.3		ug/L	1	5.0	25	10/05/12
Zinc	7440-66-6	LA-505-411	U	<5.0		ug/L	1	5.0	25	10/05/12
Calcium	7440-70-2	LA-505-411		67000		ug/L	1	49	240	10/05/12
Strontium	7440-24-6	LA-505-411		337		ug/L	1	9.0	45	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232004
SAF# W13-010
Sample ID B2M9W5

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Titanium	7440-32-6	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
Beryllium	7440-41-7	LA-505-411	U	<4.0		ug/L	1	4.0	20	10/05/12
ICPMS Prep (W)										**10/09/12**
ICP-2008 MS All possible metal										
Aluminum	7429-90-5	LA-505-412	U	<5.0		ug/L	1	5.0	50	10/10/12
Manganese	7439-96-5	LA-505-412	B	0.685		ug/L	1	0.10	1.0	10/10/12
Nickel	7440-02-0	LA-505-412		3.32		ug/L	1	0.10	1.0	10/10/12
Silver	7440-22-4	LA-505-412	U	<0.050		ug/L	1	0.050	0.50	10/10/12
Antimony	7440-36-0	LA-505-412	U	<0.30		ug/L	1	0.30	3.0	10/10/12
Barium	7440-39-3	LA-505-412		56.2		ug/L	1	0.20	2.0	10/10/12
Beryllium	7440-41-7	LA-505-412	U	<0.10		ug/L	1	0.10	1.0	10/10/12
Cadmium	7440-43-9	LA-505-412	U	<0.050		ug/L	1	0.050	0.50	10/10/12
Chromium	7440-47-3	LA-505-412		5.22		ug/L	1	0.10	1.0	10/10/12
Cobalt	7440-48-4	LA-505-412	B	0.178		ug/L	1	0.050	0.25	10/10/12
Copper	7440-50-8	LA-505-412	B	0.625		ug/L	1	0.10	1.0	10/10/12
Vanadium	7440-62-2	LA-505-412		16.2		ug/L	1	0.20	2.0	10/10/12
Zinc	7440-66-6	LA-505-412	B	1.03		ug/L	1	1.0	10	10/10/12
Lead	7439-92-1	LA-505-412	U	<0.050		ug/L	1	0.050	0.50	10/10/12
Mercury	7439-97-6	LA-505-412	U	<0.050		ug/L	1	0.050	0.20	10/10/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232004
SAF# W13-010
Sample ID B2M9W5

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Molybdenum	7439-98-7	LA-505-412		6.21		ug/L	1	0.050	0.50	10/10/12
Strontium	7440-24-6	LA-505-412		334		ug/L	1	0.10	1.0	10/10/12
Thallium	7440-28-0	LA-505-412	U	<0.050		ug/L	1	0.050	0.50	10/10/12
Tin	7440-31-5	LA-505-412	U	<0.050		ug/L	1	0.050	0.50	10/10/12
Arsenic	7440-38-2	LA-505-412		4.72		ug/L	1	0.20	2.0	10/10/12
Selenium	7782-49-2	LA-505-412	B	5.93		ug/L	1	1.0	10	10/10/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232005
SAF# W13-010
Sample ID B2M239

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for TOX (W)										**10/08/12**
Total Organic Halides										
Total Organic Halides	59473-04-0	LA-523-444	B	5.20		ug/L	1	5.0	15	10/08/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232006
SAF# W13-010
Sample ID B2M240

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for TOX (W)										**10/08/12**
Total Organic Halides										
Total Organic Halides	59473-04-0	LA-523-444	U	<5.0		ug/L	1	5.0	15	10/08/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232007
SAF# W13-010
Sample ID B2M241

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for TOX (W)										**10/08/12**
Total Organic Halides										
Total Organic Halides	59473-04-0	LA-523-444	U	<5.0		ug/L	1	5.0	15	10/08/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232008
SAF# W13-010
Sample ID B2M9W7

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for TOX (W)										**10/08/12**
Total Organic Halides										
Total Organic Halides	59473-04-0	LA-523-444	U	<5.0		ug/L	1	5.0	15	10/08/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232009
SAF# W13-010
Sample ID B2M9W8

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for TOX (W)										**10/08/12**
Total Organic Halides										
Total Organic Halides	59473-04-0	LA-523-444	U	<5.0		ug/L	1	5.0	15	10/08/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232010
SAF# W13-010
Sample ID B2M9W9

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for TOX (W)										**10/08/12**
Total Organic Halides										
Total Organic Halides	59473-04-0	LA-523-444	U	<5.0		ug/L	1	5.0	15	10/08/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232011
SAF# W13-010
Sample ID B2M8K7

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
ICPMS Prep (W)										**10/09/12**
ICP-2008 MS All possible metal										
Mercury	7439-97-6	LA-505-412	UD	<0.10		ug/L	2	0.10	0.40	10/10/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Sample # 121232012
SAF# W13-010
Sample ID B2M8K8

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
ICPMS Prep (W)										**10/09/12**
ICP-2008 MS All possible metal										
Mercury	7439-97-6	LA-505-412	UD	<0.10		ug/L	2	0.10	0.40	10/10/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the PQL(or EQL)but >= the IDL/MDL(Inorganic)
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
E - Analyte is an estimate, see comment section.
N - MS and/or MSD recovery outside control limits.

U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for 8082 (W) SPE										**10/09/12**
PCBs by EPA SW-846 Method 8082										
Aroclor-1016	12674-11-2	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1221	11104-28-2	LA-523-427	U	<0.2		ug/L	1	0.2	0.4	10/17/12
Aroclor-1232	11141-16-5	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1242	53469-21-9	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1248	12672-29-6	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1254	11097-69-1	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1260	11096-82-5	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Preparation for 8270 (W) CLE										**10/10/12**
SW-846 8270D Semivolatiles										
4-Nitrophenol	100-02-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phenol	108-95-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,2,4-Trichlorobenzene	120-82-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4-Dinitrotoluene	121-14-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,4-Dichlorobenzene	106-46-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pyrene	129-00-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Chloro-3-methylphenol	59-50-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
n-Nitroso-di-n-propylamine	621-64-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Acenaphthene	83-32-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pentachlorophenol	87-86-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Chlorophenol	95-57-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Nitroaniline	100-01-6	LA-523-456	U	<0.9		ug/L	1	0.9	2	10/16/12
4-Bromophenyl-phenylether	101-55-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4-Dimethylphenol	105-67-9	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
4-Chloroaniline	106-47-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Bis(1-Chloro-2-propyl)ether	108-60-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Bis-(2-Chloroethyl) ether	111-44-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Bis-(2-Chloroethoxy)methane	111-91-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Bis-(2-Ethylhexyl)phthalate	117-81-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Di-n-octylphthalate	117-84-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Hexachlorobenzene	118-74-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Anthracene	120-12-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4-Dichlorophenol	120-83-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Dimethylphthalate	131-11-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dibenzofuran	132-64-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(g,h,i)perylene	191-24-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Indeno(1,2,3-cd)pyrene	193-39-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(b)fluoranthene	205-99-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Fluoranthene	206-44-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(k)fluoranthene	207-08-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Acenaphthylene	208-96-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Chrysene	218-01-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(a)pyrene	50-32-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4-Dinitrophenol	51-28-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dibenzo(a,h)anthracene	53-70-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4,6-Dinitro-2-methylphenol	534-52-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,3-Dichlorobenzene	541-73-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(a)anthracene	56-55-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,6-Dinitrotoluene	606-20-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Chlorophenyl-phenylether	7005-72-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Hexachlorocyclopentadiene	77-47-4	LA-523-456	U	<0.9		ug/L	1	0.9	2	10/16/12
Isophorone	78-59-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Diethyl phthalate	84-66-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Di-n-butylphthalate	84-74-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phenanthrene	85-01-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Butylbenzylphthalate	85-68-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Fluorene	86-73-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Carbazole	86-74-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Hexachlorobutadiene	87-68-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Nitroaniline	88-74-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Nitrophenol	88-75-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Naphthalene	91-20-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Methylnaphthalene	91-57-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Chloronaphthalene	91-58-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
3,3-Dichlorobenzidine	91-94-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Methylphenol	95-48-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,2-Dichlorobenzene	95-50-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4,5-Trichlorophenol	95-95-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Nitrobenzene	98-95-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
3-Nitroaniline	99-09-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
3 & 4 Methylphenol, Total	65794-96-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Hexachloroethane	67-72-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4,6-Trichlorophenol	88-06-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzyl alcohol	100-51-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Tributyl phosphate	126-73-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Naphthylamine	91-59-8	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
Pyridine	110-86-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosopiperidine	100-75-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosomethylethylamine	10595-95-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
p-Phenylenediamine	106-50-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Picoline	109-06-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
3,3-Dimethylbenzidine	119-93-7	LA-523-456	U	<4		ug/L	1	4	6	10/16/12
Isosafrole	120-58-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phentermine	122-09-8	LA-523-456	U	<5		ug/L	1	5	9	10/16/12
1,4-Dioxane	123-91-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,4-Naphthoquinone	130-15-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
1-Naphthylamine	134-32-7	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
Aramite	140-57-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Kepone	143-50-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Hexachloropropene	1888-71-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Diallate	2303-16-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pronamide	23950-58-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Isodrin	465-73-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Chlorobenzilate	510-15-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Acetylaminofluorene	53-96-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosodiethylamine	55-18-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
3-Methylcholanthrene	56-49-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Nitroquinoline-1-oxide	56-57-5	LA-523-456	U	<0.9		ug/L	1	0.9	2	10/16/12
7,12-Dimethylbenz(a)anthracene	57-97-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,3,4,6-Tetrachlorophenol	58-90-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosomorpholine	59-89-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pentachlorobenzene	608-93-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phenacetin	62-44-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Ethyl methanesulfonate	62-50-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Aniline	62-53-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosodimethylamine	62-75-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Methyl methanesulfonate	66-27-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pentachloroethane	76-01-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pentachloronitrobenzene	82-68-8	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
2,6-Dichlorophenol	87-65-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dinoseb(..dinitromethylphenol)	88-85-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Aminobiphenyl	92-67-1	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
n-Nitrosodibutylamine	924-16-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosopyrolidine	930-55-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Safrole	94-59-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
o-Toluidine	95-53-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,2,4,5-Tetrachlorobenzene	95-94-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Acetophenone	98-86-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,3,5-Trinitrobenzene	99-35-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
2-Methyl-5-nitroaniline	99-55-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,3-Dinitrobenzene	99-65-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
O,O,O-Triethylthiophosphate	126-68-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Parathion	56-38-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dimethylaminoazobenzene	60-11-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dimethoate	60-51-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Thionazin	297-97-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Methyl parathion	298-00-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phorate	298-02-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Disulfoton	298-04-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Sulfotep	3689-24-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Famfur	52-85-7	LA-523-456	U	<5		ug/L	1	5	9	10/16/12
N-Nitrosodiphenylamin/Diphenyl	DPA+NNDPA	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Methapyrilene	91-80-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Total Petroleum Hydrocarbons (Water Prep										**10/04/12**
Extractable Diesel and Petroleum										
Diesel	TPHDIESEL	LA-523-493	U	<70		ug/L	1	70	100	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Kerosene	TPHKEROSEN	LA-523-493	U	<70		ug/L	1	70	100	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for 8082 (W) SPE										**10/09/12**
PCBs by EPA SW-846 Method 8082										
Aroclor-1016	12674-11-2	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1221	11104-28-2	LA-523-427	U	<0.2		ug/L	1	0.2	0.4	10/17/12
Aroclor-1232	11141-16-5	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1242	53469-21-9	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1248	12672-29-6	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1254	11097-69-1	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Aroclor-1260	11096-82-5	LA-523-427	U	<0.09		ug/L	1	0.09	0.2	10/17/12
Preparation for 8270 (W) CLE										**10/10/12**
SW-846 8270D Semivolatiles										
4-Nitrophenol	100-02-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phenol	108-95-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,2,4-Trichlorobenzene	120-82-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4-Dinitrotoluene	121-14-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,4-Dichlorobenzene	106-46-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pyrene	129-00-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Chloro-3-methylphenol	59-50-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
n-Nitroso-di-n-propylamine	621-64-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Acenaphthene	83-32-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pentachlorophenol	87-86-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Chlorophenol	95-57-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Nitroaniline	100-01-6	LA-523-456	U	<0.9		ug/L	1	0.9	2	10/16/12
4-Bromophenyl-phenylether	101-55-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4-Dimethylphenol	105-67-9	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
4-Chloroaniline	106-47-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Bis(1-Chloro-2-propyl)ether	108-60-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Bis-(2-Chloroethyl) ether	111-44-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Bis-(2-Chloroethoxy)methane	111-91-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Bis-(2-Ethylhexyl)phthalate	117-81-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Di-n-octylphthalate	117-84-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Hexachlorobenzene	118-74-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Anthracene	120-12-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4-Dichlorophenol	120-83-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Dimethylphthalate	131-11-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dibenzofuran	132-64-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(g,h,i)perylene	191-24-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Indeno(1,2,3-cd)pyrene	193-39-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(b)fluoranthene	205-99-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Fluoranthene	206-44-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(k)fluoranthene	207-08-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Acenaphthylene	208-96-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Chrysene	218-01-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(a)pyrene	50-32-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4-Dinitrophenol	51-28-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dibenzo(a,h)anthracene	53-70-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4,6-Dinitro-2-methylphenol	534-52-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,3-Dichlorobenzene	541-73-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzo(a)anthracene	56-55-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,6-Dinitrotoluene	606-20-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Chlorophenyl-phenylether	7005-72-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Hexachlorocyclopentadiene	77-47-4	LA-523-456	U	<0.9		ug/L	1	0.9	2	10/16/12
Isophorone	78-59-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Diethyl phthalate	84-66-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Di-n-butylphthalate	84-74-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phenanthrene	85-01-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Butylbenzylphthalate	85-68-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Fluorene	86-73-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Carbazole	86-74-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Hexachlorobutadiene	87-68-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Nitroaniline	88-74-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Nitrophenol	88-75-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Naphthalene	91-20-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Methylnaphthalene	91-57-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Chloronaphthalene	91-58-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
3,3-Dichlorobenzidine	91-94-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Methylphenol	95-48-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,2-Dichlorobenzene	95-50-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4,5-Trichlorophenol	95-95-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Nitrobenzene	98-95-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
3-Nitroaniline	99-09-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
3 & 4 Methylphenol, Total	65794-96-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Hexachloroethane	67-72-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,4,6-Trichlorophenol	88-06-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Benzyl alcohol	100-51-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Tributyl phosphate	126-73-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Naphthylamine	91-59-8	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
Pyridine	110-86-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosopiperidine	100-75-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosomethylethylamine	10595-95-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
p-Phenylenediamine	106-50-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Picoline	109-06-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
3,3-Dimethylbenzidine	119-93-7	LA-523-456	U	<4		ug/L	1	4	6	10/16/12
Isosafrole	120-58-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phentermine	122-09-8	LA-523-456	U	<5		ug/L	1	5	9	10/16/12
1,4-Dioxane	123-91-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,4-Naphthoquinone	130-15-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
1-Naphthylamine	134-32-7	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
Aramite	140-57-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Kepone	143-50-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Hexachloropropene	1888-71-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Diallate	2303-16-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pronamide	23950-58-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Isodrin	465-73-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Chlorobenzilate	510-15-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2-Acetylaminofluorene	53-96-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosodiethylamine	55-18-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
3-Methylcholanthrene	56-49-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Nitroquinoline-1-oxide	56-57-5	LA-523-456	U	<0.9		ug/L	1	0.9	2	10/16/12
7,12-Dimethylbenz(a)anthracene	57-97-6	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
2,3,4,6-Tetrachlorophenol	58-90-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosomorpholine	59-89-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pentachlorobenzene	608-93-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phenacetin	62-44-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Ethyl methanesulfonate	62-50-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Aniline	62-53-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosodimethylamine	62-75-9	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Methyl methanesulfonate	66-27-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pentachloroethane	76-01-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Pentachloronitrobenzene	82-68-8	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
2,6-Dichlorophenol	87-65-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dinoseb(..dinitromethylphenol)	88-85-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
4-Aminobiphenyl	92-67-1	LA-523-456	U	<1		ug/L	1	1	2	10/16/12
n-Nitrosodibutylamine	924-16-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
n-Nitrosopyrolidine	930-55-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Safrole	94-59-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
o-Toluidine	95-53-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,2,4,5-Tetrachlorobenzene	95-94-3	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Acetophenone	98-86-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,3,5-Trinitrobenzene	99-35-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
2-Methyl-5-nitroaniline	99-55-8	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
1,3-Dinitrobenzene	99-65-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
O,O,O-Triethylthiophosphate	126-68-1	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Parathion	56-38-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dimethylaminoazobenzene	60-11-7	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Dimethoate	60-51-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Thionazin	297-97-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Methyl parathion	298-00-0	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Phorate	298-02-2	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Disulfoton	298-04-4	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Sulfotep	3689-24-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Famfur	52-85-7	LA-523-456	U	<5		ug/L	1	5	9	10/16/12
N-Nitrosodiphenylamin/Diphenyl	DPA+NNDPA	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Methapyrilene	91-80-5	LA-523-456	U	<0.9		ug/L	1	0.9	1	10/16/12
Total Petroleum Hydrocarbons (Water Prep										**10/04/12**
Extractable Diesel and Petroleum										
Diesel	TPHDIESEL	LA-523-493	U	<70		ug/L	1	70	100	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Kerosene	TPHKEROSEN	LA-523-493	U	<70		ug/L	1	70	100	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < lowest calibration but >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - MS/MSD recovery outside control limits(GC/MS only).
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Gasoline Range Prep (W)										**10/16/12**
Gasoline Range (W)										
Gasoline	TPHGASOLIN	LA-523-443	U	<50		ug/L	1	50	200	10/16/12
Preparation for 8260B (W)										**10/11/12**
SW-846 8260B Volatiles										
1,1-Dichloroethene	75-35-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Trichloroethene	79-01-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Benzene	71-43-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Toluene	108-88-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chlorobenzene	108-90-7	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,1-Dichloroethane	75-34-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Ethylbenzene	100-41-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Styrene	100-42-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
cis-1,3-Dichloropropene	10061-01-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
trans-1,3-Dichloropropene	10061-02-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2-Dichloroethane	107-06-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methyl isobutyl ketone	108-10-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Dibromochloromethane	124-48-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < PQL (or EQL) >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - GC/MS or N - Non GC/MS - MS/MSD recovery outside control limits
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Tetrachloroethene	127-18-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Total Xylenes	1330-20-7	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Total 1,2-Dichloroethene	540-59-0	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Carbon tetrachloride	56-23-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
2-Hexanone	591-78-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Acetone	67-64-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chloroform	67-66-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,1,1-Trichloroethane	71-55-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Bromomethane	74-83-9	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chloromethane	74-87-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chloroethane	75-00-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Vinyl chloride	75-01-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methylene chloride	75-09-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Carbon disulfide	75-15-0	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Bromoform	75-25-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Bromodichloromethane	75-27-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2-Dichloropropane	78-87-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methyl ethyl ketone	78-93-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,1,2-Trichloroethane	79-00-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < PQL (or EQL) >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - GC/MS or N - Non GC/MS - MS/MSD recovery outside control limits
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
1,1,2,2-Tetrachloroethane	79-34-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1-Butanol	71-36-3	LA-523-455	U	<100		ug/L	1	100	500	10/16/12
Tetrahydrofuran	109-99-9	LA-523-455	U	<2		ug/L	1	2	10	10/16/12
Trichlorofluoromethane	75-69-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
trans-1,2-Dichloroethene	156-60-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Acetonitrile	75-05-8	LA-523-455	U	<2		ug/L	1	2	10	10/16/12
cis-1,2-Dichloroethene	156-59-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Propionitrile	107-12-0	LA-523-455	U	<2		ug/L	1	2	10	10/16/12
Isobutyl alcohol	78-83-1	LA-523-455	U	<200		ug/L	1	200	1.E3	10/16/12
Iodomethane	74-88-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,1,1,2-Tetrachloroethane	630-20-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2,3-Trichloropropane	96-18-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2-Dibromo-3-chloropropane	96-12-8	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2-Dibromoethane	106-93-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Acrolein	107-02-8	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Acrylonitrile	107-13-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Allyl chloride	107-05-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < PQL (or EQL) >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - GC/MS or N - Non GC/MS - MS/MSD recovery outside control limits
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Methylene bromide	74-95-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Dichlorodifluoromethane	75-71-8	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Ethyl methacrylate	97-63-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methacrylonitrile	126-98-7	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methyl methacrylate	80-62-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Trans-1,4-dichloro-2-butene	110-57-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Vinyl acetate	108-05-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chloroprene	126-99-8	LA-523-455	U	<1		ug/L	1	1	5	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < PQL (or EQL) >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - GC/MS or N - Non GC/MS - MS/MSD recovery outside control limits
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Gasoline Range Prep (W)										**10/16/12**
Gasoline Range (W)										
Gasoline	TPHGASOLIN	LA-523-443	U	<50		ug/L	1	50	200	10/16/12
Preparation for 8260B (W)										**10/11/12**
SW-846 8260B Volatiles										
1,1-Dichloroethene	75-35-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Trichloroethene	79-01-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Benzene	71-43-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Toluene	108-88-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chlorobenzene	108-90-7	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,1-Dichloroethane	75-34-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Ethylbenzene	100-41-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Styrene	100-42-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
cis-1,3-Dichloropropene	10061-01-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
trans-1,3-Dichloropropene	10061-02-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2-Dichloroethane	107-06-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methyl isobutyl ketone	108-10-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Dibromochloromethane	124-48-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < PQL (or EQL) >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - GC/MS or N - Non GC/MS - MS/MSD recovery outside control limits
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Tetrachloroethene	127-18-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Total Xylenes	1330-20-7	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Total 1,2-Dichloroethene	540-59-0	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Carbon tetrachloride	56-23-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
2-Hexanone	591-78-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Acetone	67-64-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chloroform	67-66-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,1,1-Trichloroethane	71-55-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Bromomethane	74-83-9	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chloromethane	74-87-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chloroethane	75-00-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Vinyl chloride	75-01-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methylene chloride	75-09-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Carbon disulfide	75-15-0	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Bromoform	75-25-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Bromodichloromethane	75-27-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2-Dichloropropane	78-87-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methyl ethyl ketone	78-93-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,1,2-Trichloroethane	79-00-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < PQL (or EQL) >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - GC/MS or N - Non GC/MS - MS/MSD recovery outside control limits
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
1,1,2,2-Tetrachloroethane	79-34-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1-Butanol	71-36-3	LA-523-455	U	<100		ug/L	1	100	500	10/16/12
Tetrahydrofuran	109-99-9	LA-523-455	U	<2		ug/L	1	2	10	10/16/12
Trichlorofluoromethane	75-69-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
trans-1,2-Dichloroethene	156-60-5	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Acetonitrile	75-05-8	LA-523-455	U	<2		ug/L	1	2	10	10/16/12
cis-1,2-Dichloroethene	156-59-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Propionitrile	107-12-0	LA-523-455	U	<2		ug/L	1	2	10	10/16/12
Isobutyl alcohol	78-83-1	LA-523-455	U	<200		ug/L	1	200	1.E3	10/16/12
Iodomethane	74-88-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,1,1,2-Tetrachloroethane	630-20-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2,3-Trichloropropane	96-18-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2-Dibromo-3-chloropropane	96-12-8	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
1,2-Dibromoethane	106-93-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Acrolein	107-02-8	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Acrylonitrile	107-13-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Allyl chloride	107-05-1	LA-523-455	U	<1		ug/L	1	1	5	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < PQL (or EQL) >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - GC/MS or N - Non GC/MS - MS/MSD recovery outside control limits
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Methylene bromide	74-95-3	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Dichlorodifluoromethane	75-71-8	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Ethyl methacrylate	97-63-2	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methacrylonitrile	126-98-7	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Methyl methacrylate	80-62-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Trans-1,4-dichloro-2-butene	110-57-6	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Vinyl acetate	108-05-4	LA-523-455	U	<1		ug/L	1	1	5	10/16/12
Chloroprene	126-99-8	LA-523-455	U	<1		ug/L	1	1	5	10/16/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte was detected in both the BLANK and SAMPLE
D - Analyte was reported at a secondary dilution factor.
E - The calibration exceeds the calibration range (GC/MS).
J - Analyte < PQL (or EQL) >= MDL.
N - Presumed evidence based on MS library search(GC/MS only)

T - GC/MS or N - Non GC/MS - MS/MSD recovery outside control limits
U - Analyzed for but not detected above limiting criteria.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Sample # 121232001
SAF# W13-010
Sample ID B2M177

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for Cyanide (W)										**10/10/12**
Cyanide (W) by Midi/Spectrophotometer										
Cyanide	57-12-5	LA-695-402		84.4		ug/L	1	4.0	20	10/10/12
										10/15/12
Total Alkalinity as mg/L CaCO3 (Water)										
Total Alkalinity as CaCO3	ALKALINITY	LA-531-411		110		mg/L	1	1	10	10/15/12
										10/05/12
Total Organic Carbon										
Total Organic Carbon	TOC	LA-344-406		0.333		mg/L	1	0.10	0.30	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the RDL but >= the IDL/MDL.
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
N - MS and/or MSD sample recovery outside control limits.
U - Analyzed for but not detected above limiting criteria.

N - Spike Recovery is Outside Control Limits.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Sample # 121232002
SAF# W13-010
Sample ID B2M9W6

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
Preparation for Cyanide (W)										**10/10/12**
Cyanide (W) by Midi/Spectrophotometer										
Cyanide	57-12-5	LA-695-402		22.8		ug/L	1	4.0	20	10/10/12
										10/15/12
Total Alkalinity as mg/L CaCO3 (Water)										
Total Alkalinity as CaCO3	ALKALINITY	LA-531-411		110		mg/L	1	1	10	10/15/12
										10/05/12
Total Organic Carbon										
Total Organic Carbon	TOC	LA-344-406	B	0.249		mg/L	1	0.10	0.30	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the RDL but >= the IDL/MDL.
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
N - MS and/or MSD sample recovery outside control limits.
U - Analyzed for but not detected above limiting criteria.

N - Spike Recovery is Outside Control Limits.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Sample # 121232005
SAF# W13-010
Sample ID B2M239

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
										10/05/12
Total Organic Carbon										
Total Organic Carbon	TOC	LA-344-406		0.320		mg/L	1	0.10	0.30	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the RDL but >= the IDL/MDL.
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
N - MS and/or MSD sample recovery outside control limits.
U - Analyzed for but not detected above limiting criteria.

N - Spike Recovery is Outside Control Limits.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Sample # 121232006
SAF# W13-010
Sample ID B2M240

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
										10/05/12
Total Organic Carbon										
Total Organic Carbon	TOC	LA-344-406		0.332		mg/L	1	0.10	0.30	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the RDL but >= the IDL/MDL.
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
N - MS and/or MSD sample recovery outside control limits.
U - Analyzed for but not detected above limiting criteria.

N - Spike Recovery is Outside Control Limits.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Sample # 121232007
SAF# W13-010
Sample ID B2M241

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
										10/05/12
Total Organic Carbon										
Total Organic Carbon	TOC	LA-344-406		0.325		mg/L	1	0.10	0.30	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the RDL but >= the IDL/MDL.
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
N - MS and/or MSD sample recovery outside control limits.
U - Analyzed for but not detected above limiting criteria.

N - Spike Recovery is Outside Control Limits.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Sample # 121232008
SAF# W13-010
Sample ID B2M9W7

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
										10/05/12
Total Organic Carbon										
Total Organic Carbon	TOC	LA-344-406	B	0.245		mg/L	1	0.10	0.30	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the RDL but >= the IDL/MDL.
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
N - MS and/or MSD sample recovery outside control limits.
U - Analyzed for but not detected above limiting criteria.

N - Spike Recovery is Outside Control Limits.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Sample # 121232009
SAF# W13-010
Sample ID B2M9W8

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
										10/05/12
Total Organic Carbon										
Total Organic Carbon	TOC	LA-344-406	B	0.246		mg/L	1	0.10	0.30	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the RDL but >= the IDL/MDL.
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
N - MS and/or MSD sample recovery outside control limits.
U - Analyzed for but not detected above limiting criteria.

N - Spike Recovery is Outside Control Limits.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Sample # 121232010
SAF# W13-010
Sample ID B2M9W9

Matrix WATER
Sampled 10/03/12
Received 10/03/12

Test Performed	CAS #	Method	RQ	Result	TP Err	Units	DF	MDL	PQL	Analyzed
										10/05/12
Total Organic Carbon										
Total Organic Carbon	TOC	LA-344-406	B	0.251		mg/L	1	0.10	0.30	10/05/12

MDL = Minimum Detection Limit
RQ = Result Qualifier
TP Err = Total Propagated Error
DF = Dilution Factor
+ - Indicates more than nine qualifier

B - Analyte < the RDL but >= the IDL/MDL.
C - Analyte was found in the Associated Blank. (Inorganic)
D - Analyte was reported at a secondary dilution factor.
N - MS and/or MSD sample recovery outside control limits.
U - Analyzed for but not detected above limiting criteria.

N - Spike Recovery is Outside Control Limits.
X,Y or Z - See comment detail and/or narrative.
PQL is equivalent to Estimated Quantitation Limit (EQL)
o - LCS recovery outside established laboratory acceptance limits.

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analytical Batch 208205 (QC Batch: 208198) **Test** ICP-6010 - All possible metals
Associated Samples 121232001, 121232002, 121232003, 121232004

Analyte	CAS #	Original Found	QC Found	Units	% Recov Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82193**						
Iron	7439-89-6		<19	ug/L				U	10/05/12
Magnesium	7439-95-4		<4.0	ug/L				U	10/05/12
Manganese	7439-96-5		<4.0	ug/L				U	10/05/12
Nickel	7440-02-0		<4.0	ug/L				U	10/05/12
Potassium	7440-09-7		<76	ug/L				U	10/05/12
Silver	7440-22-4		<4.0	ug/L				U	10/05/12
Sodium	7440-23-5		<10	ug/L				U	10/05/12
Antimony	7440-36-0		<36	ug/L				U	10/05/12
Barium	7440-39-3		<4.0	ug/L				U	10/05/12
Cadmium	7440-43-9		<4.0	ug/L				U	10/05/12
Chromium	7440-47-3		<5.0	ug/L				U	10/05/12
Cobalt	7440-48-4		<4.0	ug/L				U	10/05/12
Copper	7440-50-8		<4.0	ug/L				U	10/05/12
Vanadium	7440-62-2		<5.0	ug/L				U	10/05/12
Zinc	7440-66-6		<5.0	ug/L				U	10/05/12
Calcium	7440-70-2		<49	ug/L				U	10/05/12
Strontium	7440-24-6		<9.0	ug/L				U	10/05/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Titanium	7440-32-6		<4.0	ug/L					U	10/05/12
Beryllium	7440-41-7		<4.0	ug/L					U	10/05/12
LCS			**QC Sample #82195**							
Iron	7439-89-6		1010	ug/L	100.6	80 - 120				10/05/12
Magnesium	7439-95-4		10300	ug/L	103	80 - 120				10/05/12
Manganese	7439-96-5		1030	ug/L	102.7	80 - 120				10/05/12
Nickel	7440-02-0		997	ug/L	99.7	80 - 120				10/05/12
Potassium	7440-09-7		10900	ug/L	108.6	80 - 120				10/05/12
Silver	7440-22-4		1020	ug/L	102.2	80 - 120				10/05/12
Sodium	7440-23-5		10500	ug/L	105	80 - 120				10/05/12
Antimony	7440-36-0		1040	ug/L	103.9	80 - 120				10/05/12
Barium	7440-39-3		1040	ug/L	104.5	80 - 120				10/05/12
Cadmium	7440-43-9		1010	ug/L	100.9	80 - 120				10/05/12
Chromium	7440-47-3		1010	ug/L	101.2	80 - 120				10/05/12
Cobalt	7440-48-4		991	ug/L	99.1	80 - 120				10/05/12
Copper	7440-50-8		1030	ug/L	102.8	80 - 120				10/05/12
Vanadium	7440-62-2		1010	ug/L	100.6	80 - 120				10/05/12
Zinc	7440-66-6		1030	ug/L	103.3	80 - 120				10/05/12
Calcium	7440-70-2		20700	ug/L	103.7	80 - 120				10/05/12
Strontium	7440-24-6		988	ug/L	98.8	80 - 120				10/05/12
Titanium	7440-32-6		1030	ug/L	103.4	80 - 120				10/05/12
Beryllium	7440-41-7		1010	ug/L	101.4	80 - 120				10/05/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
MS			**QC Sample #82196**							
			Original 121223007							
Iron	7439-89-6		1010	ug/L	101.2	75 - 125				10/05/12
Magnesium	7439-95-4		10400	ug/L	104.5	75 - 125				10/05/12
Manganese	7439-96-5		1030	ug/L	103.4	75 - 125				10/05/12
Nickel	7440-02-0		988	ug/L	98.8	75 - 125				10/05/12
Potassium	7440-09-7		11000	ug/L	110	75 - 125				10/05/12
Silver	7440-22-4		1020	ug/L	102	75 - 125				10/05/12
Sodium	7440-23-5		9960	ug/L	99.6	75 - 125				10/05/12
Antimony	7440-36-0		1030	ug/L	103.4	75 - 125				10/05/12
Barium	7440-39-3		1030	ug/L	103.4	75 - 125				10/05/12
Cadmium	7440-43-9		1010	ug/L	101.2	75 - 125				10/05/12
Chromium	7440-47-3		1020	ug/L	102.4	75 - 125				10/05/12
Cobalt	7440-48-4		984	ug/L	98.4	75 - 125				10/05/12
Copper	7440-50-8		1020	ug/L	101.5	75 - 125				10/05/12
Vanadium	7440-62-2		1010	ug/L	101.5	75 - 125				10/05/12
Zinc	7440-66-6		1040	ug/L	103.7	75 - 125				10/05/12
Calcium	7440-70-2		21800	ug/L	108.8	75 - 125				10/05/12
Strontium	7440-24-6		1000	ug/L	100.3	75 - 125				10/05/12
Titanium	7440-32-6		1030	ug/L	103.1	75 - 125				10/05/12
Beryllium	7440-41-7		1030	ug/L	102.6	75 - 125				10/05/12
MSD			**QC Sample #82197**							
			Original 121223007						**Paired 82196**	
Iron	7439-89-6		1000	ug/L	100.5	75 - 125	0.70	20		10/05/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Magnesium	7439-95-4		10300	ug/L	103.3	75 - 125	0.50	20		10/05/12
Manganese	7439-96-5		1030	ug/L	102.9	75 - 125	0.50	20		10/05/12
Nickel	7440-02-0		984	ug/L	98.4	75 - 125	0.40	20		10/05/12
Potassium	7440-09-7		10900	ug/L	108.5	75 - 125	0.80	20		10/05/12
Silver	7440-22-4		1020	ug/L	101.7	75 - 125	0.30	20		10/05/12
Sodium	7440-23-5		9930	ug/L	99.3	75 - 125	0.10	20		10/05/12
Antimony	7440-36-0		1050	ug/L	104.6	75 - 125	1.20	20		10/05/12
Barium	7440-39-3		1030	ug/L	103.5	75 - 125	0.10	20		10/05/12
Cadmium	7440-43-9		1010	ug/L	101.3	75 - 125	0.10	20		10/05/12
Chromium	7440-47-3		1020	ug/L	102.1	75 - 125	0.30	20		10/05/12
Cobalt	7440-48-4		981	ug/L	98.1	75 - 125	0.30	20		10/05/12
Copper	7440-50-8		1010	ug/L	101.4	75 - 125	0.10	20		10/05/12
Vanadium	7440-62-2		1010	ug/L	101	75 - 125	0.50	20		10/05/12
Zinc	7440-66-6		1040	ug/L	103.8	75 - 125	0.10	20		10/05/12
Calcium	7440-70-2		21200	ug/L	105.8	75 - 125	0.90	20		10/05/12
Strontium	7440-24-6		999	ug/L	99.9	75 - 125	0.30	20		10/05/12
Titanium	7440-32-6		1030	ug/L	102.9	75 - 125	0.20	20		10/05/12
Beryllium	7440-41-7		1020	ug/L	102.5	75 - 125	0.10	20		10/05/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analytical Batch 208248 (QC Batch: 208247) **Test** Chemical Oxygen Demand
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82359**							
Chemical Oxygen Demand	COD		<10	mg/L					U	10/04/12
LCS			**QC Sample #82360**							
Chemical Oxygen Demand	COD		100	mg/L	100.4	80 - 120				10/04/12
MS			**QC Sample #82363**							
			Original 121215007							
Chemical Oxygen Demand	COD		267	mg/L	106.8	75 - 125				10/04/12
MSD			**QC Sample #82364**							
			Original 121215007					**Paired 82363**		
Chemical Oxygen Demand	COD		260	mg/L	104	75 - 125	2.70	20		10/04/12

* - QC result out of range n/a - Not Applicable

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analytical Batch 208258 (QC Batch: 208245) **Test** Extractable Diesel and Petroleum
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82350**							
Diesel	TPHDIESEL		<80	ug/L					U	10/05/12
Kerosene	TPHKEROSE		<80	ug/L					U	10/05/12
LCS			**QC Sample #82351**							
Diesel	TPHDIESEL		2800	ug/L	111.6	65 - 128				10/05/12
MS			**QC Sample #82352**							
			Original 121223013							
Diesel	TPHDIESEL		2400	ug/L	102.2	73 - 123				10/05/12
MSD			**QC Sample #82353**							
			Original 121223013					**Paired**	**82352**	
Diesel	TPHDIESEL		2500	ug/L	105.7	73 - 123	3.40	20		10/05/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Analytical Batch 208268 (QC Batch: 208268) **Test** Total Organic Carbon
Associated Samples 121232001, 121232002, 121232005, 121232006, 121232007, 121232008, 121232009, 121232010

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82439**							
Total Organic Carbon	TOC		<0.045	mg/L					U	10/05/12
LCS			**QC Sample #82440**							
Total Organic Carbon	TOC		2.20	mg/L	109.8	80 - 120				10/05/12
MS			**QC Sample #82444**							
			Original 121230004							
Total Organic Carbon	TOC		2.21	mg/L	110.3	75 - 125				10/05/12
MSD			**QC Sample #82445**							
			Original 121230004					**Paired 82444**		
Total Organic Carbon	TOC		2.20	mg/L	109.9	75 - 125	0.30	20		10/05/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analytical Batch 208292 (QC Batch: 208257) **Test** ICP-2008 MS All possible metal
Associated Samples 121232003

Analyte	CAS #	Original Found	QC Found	Units	% Recov Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82385**						
Aluminum	7429-90-5		16.0	ug/L				B	10/09/12
Manganese	7439-96-5		<0.10	ug/L				U	10/09/12
Nickel	7440-02-0		<0.10	ug/L				U	10/09/12
Silver	7440-22-4		<0.050	ug/L				U	10/09/12
Antimony	7440-36-0		<0.30	ug/L				U	10/09/12
Barium	7440-39-3		<0.20	ug/L				U	10/09/12
Beryllium	7440-41-7		<0.10	ug/L				U	10/09/12
Cadmium	7440-43-9		<0.050	ug/L				U	10/09/12
Chromium	7440-47-3		<0.10	ug/L				U	10/09/12
Cobalt	7440-48-4		<0.050	ug/L				U	10/09/12
Copper	7440-50-8		0.155	ug/L				B	10/09/12
Vanadium	7440-62-2		0.315	ug/L				B	10/09/12
Zinc	7440-66-6		<1.0	ug/L				U	10/09/12
Lead	7439-92-1		<0.050	ug/L				U	10/09/12
Molybdenum	7439-98-7		<0.050	ug/L				U	10/09/12
Strontium	7440-24-6		0.242	ug/L				B	10/09/12
Thallium	7440-28-0		<0.050	ug/L				U	10/09/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Tin	7440-31-5		<0.050	ug/L					U	10/09/12
Arsenic	7440-38-2		<0.20	ug/L					U	10/09/12
Selenium	7782-49-2		<1.0	ug/L					U	10/09/12
LCS			**QC Sample #82386**							
Aluminum	7429-90-5		448	ug/L	112	85 - 115				10/09/12
Manganese	7439-96-5		42.3	ug/L	105.7	85 - 115				10/09/12
Nickel	7440-02-0		42.4	ug/L	105.9	85 - 115				10/09/12
Silver	7440-22-4		41.8	ug/L	104.6	85 - 115				10/09/12
Antimony	7440-36-0		41.1	ug/L	102.6	85 - 115				10/09/12
Barium	7440-39-3		43.4	ug/L	108.4	85 - 115				10/09/12
Beryllium	7440-41-7		44.0	ug/L	110	85 - 115				10/09/12
Cadmium	7440-43-9		41.7	ug/L	104.4	85 - 115				10/09/12
Chromium	7440-47-3		43.1	ug/L	107.8	85 - 115				10/09/12
Cobalt	7440-48-4		42.4	ug/L	106.1	85 - 115				10/09/12
Copper	7440-50-8		42.7	ug/L	106.8	85 - 115				10/09/12
Vanadium	7440-62-2		43.3	ug/L	108.3	85 - 115				10/09/12
Zinc	7440-66-6		39.6	ug/L	99	85 - 115				10/09/12
Lead	7439-92-1		44.5	ug/L	111.3	85 - 115				10/09/12
Molybdenum	7439-98-7		42.2	ug/L	105.4	85 - 115				10/09/12
Strontium	7440-24-6		424	ug/L	106	85 - 115				10/09/12
Thallium	7440-28-0		44.2	ug/L	110.4	85 - 115				10/09/12
Tin	7440-31-5		42.1	ug/L	105.2	85 - 115				10/09/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Arsenic	7440-38-2		40.5	ug/L	101.2	85 - 115				10/09/12
Selenium	7782-49-2		37.7	ug/L	94.2	85 - 115				10/09/12
MS			**QC Sample #82387**							
			Original 121223010							
Aluminum	7429-90-5		453	ug/L	113.1	70 - 130				10/09/12
Manganese	7439-96-5		42.1	ug/L	105.3	70 - 130				10/09/12
Nickel	7440-02-0		40.8	ug/L	102.1	70 - 130				10/09/12
Silver	7440-22-4		39.8	ug/L	99.6	70 - 130				10/09/12
Antimony	7440-36-0		42.7	ug/L	106.7	70 - 130				10/09/12
Barium	7440-39-3		42.8	ug/L	107.1	70 - 130				10/09/12
Beryllium	7440-41-7		44.8	ug/L	112.1	70 - 130				10/09/12
Cadmium	7440-43-9		41.8	ug/L	104.6	70 - 130				10/09/12
Chromium	7440-47-3		43.4	ug/L	108.4	70 - 130				10/09/12
Cobalt	7440-48-4		41.7	ug/L	104.2	70 - 130				10/09/12
Copper	7440-50-8		39.8	ug/L	99.4	70 - 130				10/09/12
Vanadium	7440-62-2		44.4	ug/L	111	70 - 130				10/09/12
Zinc	7440-66-6		38.0	ug/L	95	70 - 130				10/09/12
Lead	7439-92-1		46.4	ug/L	115.9	70 - 130				10/09/12
Molybdenum	7439-98-7		44.7	ug/L	111.7	70 - 130				10/09/12
Strontium	7440-24-6		427	ug/L	106.7	70 - 130				10/09/12
Thallium	7440-28-0		46.8	ug/L	116.9	70 - 130				10/09/12
Tin	7440-31-5		43.4	ug/L	108.5	70 - 130				10/09/12
Arsenic	7440-38-2		41.6	ug/L	103.9	70 - 130				10/09/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Selenium	7782-49-2		38.2	ug/L	95.5	70 - 130				10/09/12
MSD			**QC Sample #82388**							
			Original 121223010						**Paired 82387**	
Aluminum	7429-90-5		453	ug/L	113.2	70 - 130	0.00	20		10/09/12
Manganese	7439-96-5		42.3	ug/L	105.7	70 - 130	0.40	20		10/09/12
Nickel	7440-02-0		40.9	ug/L	102.3	70 - 130	0.20	20		10/09/12
Silver	7440-22-4		40.0	ug/L	100	70 - 130	0.40	20		10/09/12
Antimony	7440-36-0		42.4	ug/L	106.1	70 - 130	0.60	20		10/09/12
Barium	7440-39-3		42.5	ug/L	106.4	70 - 130	0.20	20		10/09/12
Beryllium	7440-41-7		46.4	ug/L	116.1	70 - 130	3.50	20		10/09/12
Cadmium	7440-43-9		41.4	ug/L	103.6	70 - 130	0.90	20		10/09/12
Chromium	7440-47-3		43.3	ug/L	108.3	70 - 130	0.10	20		10/09/12
Cobalt	7440-48-4		41.5	ug/L	103.7	70 - 130	0.50	20		10/09/12
Copper	7440-50-8		39.8	ug/L	99.4	70 - 130	0.00	20		10/09/12
Vanadium	7440-62-2		44.7	ug/L	111.7	70 - 130	0.40	20		10/09/12
Zinc	7440-66-6		38.6	ug/L	96.6	70 - 130	1.70	20		10/09/12
Lead	7439-92-1		46.1	ug/L	115.3	70 - 130	0.60	20		10/09/12
Molybdenum	7439-98-7		44.6	ug/L	111.6	70 - 130	0.10	20		10/09/12
Strontium	7440-24-6		434	ug/L	108.5	70 - 130	1.10	20		10/09/12
Thallium	7440-28-0		46.5	ug/L	116.2	70 - 130	0.60	20		10/09/12
Tin	7440-31-5		43.2	ug/L	108	70 - 130	0.50	20		10/09/12
Arsenic	7440-38-2		41.8	ug/L	104.5	70 - 130	0.60	20		10/09/12
Selenium	7782-49-2		38.3	ug/L	95.7	70 - 130	0.20	20		10/09/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analytical Batch 208303 (QC Batch: 208289)
Associated Samples 121232004, 121232011, 121232012

Test ICP-2008 MS All possible metal

Analyte	CAS #	Original Found	QC Found	Units	% Recov Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82511**						
Aluminum	7429-90-5		<5.0	ug/L				U	10/10/12
Manganese	7439-96-5		<0.10	ug/L				U	10/10/12
Nickel	7440-02-0		<0.10	ug/L				U	10/10/12
Silver	7440-22-4		<0.050	ug/L				U	10/10/12
Antimony	7440-36-0		<0.30	ug/L				U	10/10/12
Barium	7440-39-3		<0.20	ug/L				U	10/10/12
Beryllium	7440-41-7		<0.10	ug/L				U	10/10/12
Cadmium	7440-43-9		<0.050	ug/L				U	10/10/12
Chromium	7440-47-3		<0.10	ug/L				U	10/10/12
Cobalt	7440-48-4		<0.050	ug/L				U	10/10/12
Copper	7440-50-8		<0.10	ug/L				U	10/10/12
Vanadium	7440-62-2		0.576	ug/L				B	10/10/12
Zinc	7440-66-6		<1.0	ug/L				U	10/10/12
Lead	7439-92-1		<0.050	ug/L				U	10/10/12
Mercury	7439-97-6		<0.050	ug/L				U	10/10/12
Molybdenum	7439-98-7		<0.050	ug/L				U	10/10/12
Strontium	7440-24-6		<0.10	ug/L				U	10/10/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Thallium	7440-28-0		<0.050	ug/L					U	10/10/12
Tin	7440-31-5		<0.050	ug/L					U	10/10/12
Arsenic	7440-38-2		<0.20	ug/L					U	10/10/12
Selenium	7782-49-2		<1.0	ug/L					U	10/10/12
LCS			**QC Sample #82512**							
Aluminum	7429-90-5		415	ug/L	103.8	85 - 115				10/10/12
Manganese	7439-96-5		41.7	ug/L	104.2	85 - 115				10/10/12
Nickel	7440-02-0		41.3	ug/L	103.2	85 - 115				10/10/12
Silver	7440-22-4		41.7	ug/L	104.4	85 - 115				10/10/12
Antimony	7440-36-0		40.0	ug/L	100.1	85 - 115				10/10/12
Barium	7440-39-3		42.2	ug/L	105.5	85 - 115				10/10/12
Beryllium	7440-41-7		41.7	ug/L	104.2	85 - 115				10/10/12
Cadmium	7440-43-9		40.7	ug/L	101.8	85 - 115				10/10/12
Chromium	7440-47-3		41.5	ug/L	103.8	85 - 115				10/10/12
Cobalt	7440-48-4		41.5	ug/L	103.7	85 - 115				10/10/12
Copper	7440-50-8		41.3	ug/L	103.4	85 - 115				10/10/12
Vanadium	7440-62-2		42.2	ug/L	105.6	85 - 115				10/10/12
Zinc	7440-66-6		37.3	ug/L	93.3	85 - 115				10/10/12
Lead	7439-92-1		43.8	ug/L	109.6	85 - 115				10/10/12
Mercury	7439-97-6		1.72	ug/L	86	85 - 115				10/10/12
Molybdenum	7439-98-7		41.7	ug/L	104.4	85 - 115				10/10/12
Strontium	7440-24-6		419	ug/L	104.8	85 - 115				10/10/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Thallium	7440-28-0		43.2	ug/L	108.1	85 - 115				10/10/12
Tin	7440-31-5		41.7	ug/L	104.4	85 - 115				10/10/12
Arsenic	7440-38-2		39.9	ug/L	99.7	85 - 115				10/10/12
Selenium	7782-49-2		36.9	ug/L	92.3	85 - 115				10/10/12
MS			**QC Sample #82513**							
			Original 121232004							
Aluminum	7429-90-5	<5.0	402	ug/L	100.5	70 - 130				10/10/12
Manganese	7439-96-5	0.685	39.7	ug/L	99.2	70 - 130				10/10/12
Nickel	7440-02-0	3.32	37.9	ug/L	94.8	70 - 130				10/10/12
Silver	7440-22-4	<0.050	37.6	ug/L	94	70 - 130				10/10/12
Antimony	7440-36-0	<0.30	39.8	ug/L	99.5	70 - 130				10/10/12
Barium	7440-39-3	56.2	41.0	ug/L	102.6	70 - 130				10/10/12
Beryllium	7440-41-7	<0.10	40.2	ug/L	100.5	70 - 130				10/10/12
Cadmium	7440-43-9	<0.050	38.7	ug/L	96.8	70 - 130				10/10/12
Chromium	7440-47-3	5.22	39.6	ug/L	99.1	70 - 130				10/10/12
Cobalt	7440-48-4	0.178	38.8	ug/L	96.9	70 - 130				10/10/12
Copper	7440-50-8	0.625	35.7	ug/L	89.2	70 - 130				10/10/12
Vanadium	7440-62-2	16.2	41.6	ug/L	104	70 - 130				10/10/12
Zinc	7440-66-6	1.03	33.8	ug/L	84.6	70 - 130				10/10/12
Lead	7439-92-1	<0.050	43.6	ug/L	108.9	70 - 130				10/10/12
Mercury	7439-97-6	<0.050	1.80	ug/L	90.2	70 - 130				10/10/12
Molybdenum	7439-98-7	6.21	41.9	ug/L	104.8	70 - 130				10/10/12
Strontium	7440-24-6	334	409	ug/L	102.2	70 - 130				10/10/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Thallium	7440-28-0	<0.050	43.7	ug/L	109.2	70 - 130				10/10/12
Tin	7440-31-5	<0.050	40.6	ug/L	101.4	70 - 130				10/10/12
Arsenic	7440-38-2	4.72	39.6	ug/L	99.1	70 - 130				10/10/12
Selenium	7782-49-2	5.93	36.7	ug/L	91.7	70 - 130				10/10/12
MSD			**QC Sample #82514**							
			Original 121232004					**Paired**	**82513**	
Aluminum	7429-90-5	<5.0	392	ug/L	97.9	70 - 130	2.70	20		10/10/12
Manganese	7439-96-5	0.685	38.7	ug/L	96.7	70 - 130	2.50	20		10/10/12
Nickel	7440-02-0	3.32	37.4	ug/L	93.4	70 - 130	1.40	20		10/10/12
Silver	7440-22-4	<0.050	38.0	ug/L	95	70 - 130	1.10	20		10/10/12
Antimony	7440-36-0	<0.30	39.8	ug/L	99.6	70 - 130	0.10	20		10/10/12
Barium	7440-39-3	56.2	42.2	ug/L	105.5	70 - 130	1.20	20		10/10/12
Beryllium	7440-41-7	<0.10	39.3	ug/L	98.2	70 - 130	2.40	20		10/10/12
Cadmium	7440-43-9	<0.050	39.1	ug/L	97.7	70 - 130	1.00	20		10/10/12
Chromium	7440-47-3	5.22	38.6	ug/L	96.5	70 - 130	2.30	20		10/10/12
Cobalt	7440-48-4	0.178	37.8	ug/L	94.5	70 - 130	2.50	20		10/10/12
Copper	7440-50-8	0.625	35.1	ug/L	87.7	70 - 130	1.70	20		10/10/12
Vanadium	7440-62-2	16.2	40.3	ug/L	100.8	70 - 130	2.20	20		10/10/12
Zinc	7440-66-6	1.03	33.0	ug/L	82.6	70 - 130	2.30	20		10/10/12
Lead	7439-92-1	<0.050	43.4	ug/L	108.5	70 - 130	0.30	20		10/10/12
Mercury	7439-97-6	<0.050	1.80	ug/L	89.8	70 - 130	0.40	20		10/10/12
Molybdenum	7439-98-7	6.21	42.0	ug/L	105	70 - 130	0.10	20		10/10/12
Strontium	7440-24-6	334	418	ug/L	104.5	70 - 130	1.20	20		10/10/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Thallium	7440-28-0	<0.050	43.3	ug/L	108.4	70 - 130	0.80	20		10/10/12
Tin	7440-31-5	<0.050	40.5	ug/L	101.3	70 - 130	0.10	20		10/10/12
Arsenic	7440-38-2	4.72	39.9	ug/L	99.7	70 - 130	0.50	20		10/10/12
Selenium	7782-49-2	5.93	36.7	ug/L	91.7	70 - 130	0.00	20		10/10/12

* - QC result out of range n/a - Not Applicable

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analytical Batch 208457 (QC Batch: 208456) **Test** SW-846 8260B Volatiles
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82689**						
1,1-Dichloroethene	75-35-4		<1	ug/L				U	10/16/12
Trichloroethene	79-01-6		<1	ug/L				U	10/16/12
Benzene	71-43-2		<1	ug/L				U	10/16/12
Toluene	108-88-3		<1	ug/L				U	10/16/12
Chlorobenzene	108-90-7		<1	ug/L				U	10/16/12
1,1-Dichloroethane	75-34-3		<1	ug/L				U	10/16/12
Ethylbenzene	100-41-4		<1	ug/L				U	10/16/12
Styrene	100-42-5		<1	ug/L				U	10/16/12
cis-1,3-Dichloropropene	10061-01-5		<1	ug/L				U	10/16/12
trans-1,3-Dichloropropene	10061-02-6		<1	ug/L				U	10/16/12
1,2-Dichloroethane	107-06-2		<1	ug/L				U	10/16/12
Methyl isobutyl ketone	108-10-1		<1	ug/L				U	10/16/12
Dibromochloromethane	124-48-1		<1	ug/L				U	10/16/12
Tetrachloroethene	127-18-4		<1	ug/L				U	10/16/12
Total Xylenes	1330-20-7		<1	ug/L				U	10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Total 1,2-Dichloroethene	540-59-0		<1	ug/L					U	10/16/12
Carbon tetrachloride	56-23-5		<1	ug/L					U	10/16/12
2-Hexanone	591-78-6		<1	ug/L					U	10/16/12
Acetone	67-64-1		<1	ug/L					U	10/16/12
Chloroform	67-66-3		<1	ug/L					U	10/16/12
1,1,1-Trichloroethane	71-55-6		<1	ug/L					U	10/16/12
Bromomethane	74-83-9		<1	ug/L					U	10/16/12
Chloromethane	74-87-3		<1	ug/L					U	10/16/12
Chloroethane	75-00-3		<1	ug/L					U	10/16/12
Vinyl chloride	75-01-4		<1	ug/L					U	10/16/12
Methylene chloride	75-09-2		<1	ug/L					U	10/16/12
Carbon disulfide	75-15-0		<1	ug/L					U	10/16/12
Bromoform	75-25-2		<1	ug/L					U	10/16/12
Bromodichloromethane	75-27-4		<1	ug/L					U	10/16/12
1,2-Dichloropropane	78-87-5		<1	ug/L					U	10/16/12
Methyl ethyl ketone	78-93-3		<1	ug/L					U	10/16/12
1,1,2-Trichloroethane	79-00-5		<1	ug/L					U	10/16/12
1,1,2,2-Tetrachloroethane	79-34-5		<1	ug/L					U	10/16/12
1-Butanol	71-36-3		<100	ug/L					U	10/16/12
Tetrahydrofuran	109-99-9		<2	ug/L					U	10/16/12
Trichlorofluoromethane	75-69-4		<1	ug/L					U	10/16/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
trans-1,2-Dichloroethene	156-60-5		<1	ug/L					U	10/16/12
Acetonitrile	75-05-8		<2	ug/L					U	10/16/12
cis-1,2-Dichloroethene	156-59-2		<1	ug/L					U	10/16/12
Propionitrile	107-12-0		<2	ug/L					U	10/16/12
Isobutyl alcohol	78-83-1		<200	ug/L					U	10/16/12
Iodomethane	74-88-4		<1	ug/L					U	10/16/12
1,1,1,2-Tetrachloroethane	630-20-6		<1	ug/L					U	10/16/12
1,2,3-Trichloropropane	96-18-4		<1	ug/L					U	10/16/12
1,2-Dibromo-3-chloropropane	96-12-8		<1	ug/L					U	10/16/12
1,2-Dibromoethane	106-93-4		<1	ug/L					U	10/16/12
Acrolein	107-02-8		<1	ug/L					U	10/16/12
Acrylonitrile	107-13-1		<1	ug/L					U	10/16/12
Allyl chloride	107-05-1		<1	ug/L					U	10/16/12
Methylene bromide	74-95-3		<1	ug/L					U	10/16/12
Dichlorodifluoromethane	75-71-8		<1	ug/L					U	10/16/12
Ethyl methacrylate	97-63-2		<1	ug/L					U	10/16/12
Methacrylonitrile	126-98-7		<1	ug/L					U	10/16/12
Methyl methacrylate	80-62-6		<1	ug/L					U	10/16/12
Trans-1,4-dichloro-2-butene	110-57-6		<1	ug/L					U	10/16/12
Vinyl acetate	108-05-4		<1	ug/L					U	10/16/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Chloroprene	126-99-8		<1	ug/L					U	10/16/12
LCS			**QC Sample #82690**							
1,1-Dichloroethene	75-35-4		25	ug/L	99.9	75 - 125				10/16/12
Trichloroethene	79-01-6		24	ug/L	97.2	75 - 125				10/16/12
Benzene	71-43-2		26	ug/L	103.8	75 - 125				10/16/12
Toluene	108-88-3		25	ug/L	99.4	75 - 125				10/16/12
Chlorobenzene	108-90-7		25	ug/L	101.5	75 - 125				10/16/12
1,1-Dichloroethane	75-34-3		25	ug/L	100.4	75 - 125				10/16/12
Ethylbenzene	100-41-4		26	ug/L	102.4	75 - 125				10/16/12
Styrene	100-42-5		27	ug/L	109.9	75 - 125				10/16/12
trans-1,3-Dichloropropene	10061-02-6		27	ug/L	106.9	75 - 125				10/16/12
1,2-Dichloroethane	107-06-2		28	ug/L	111	75 - 125				10/16/12
1,1,1-Trichloroethane	71-55-6		26	ug/L	103.1	75 - 125				10/16/12
Dibromochloromethane	124-48-1		28	ug/L	111.1	75 - 125				10/16/12
Carbon disulfide	75-15-0		25	ug/L	99.2	75 - 125				10/16/12
Bromoform	75-25-2		31	ug/L	125	75 - 125				10/16/12
Bromodichloromethane	75-27-4		27	ug/L	107.9	75 - 125				10/16/12
1,2-Dichloropropane	78-87-5		27	ug/L	106.4	75 - 125				10/16/12
1,1,2-Trichloroethane	79-00-5		28	ug/L	110.6	75 - 125				10/16/12
1,1,2,2-Tetrachloroethane	79-34-5		29	ug/L	116.2	75 - 125				10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
trans-1,2-Dichloroethene	156-60-5		25	ug/L	100.8	75 - 125				10/16/12
cis-1,2-Dichloroethene	156-59-2		25	ug/L	98.2	75 - 125				10/16/12
MS			**QC Sample #82691**							
			Original 121223013							
1,1-Dichloroethene	75-35-4		25	ug/L	100.2	75 - 125				10/16/12
Trichloroethene	79-01-6		25	ug/L	98.9	75 - 125				10/16/12
Benzene	71-43-2		26	ug/L	105.3	75 - 125				10/16/12
Toluene	108-88-3		25	ug/L	101.8	75 - 125				10/16/12
Chlorobenzene	108-90-7		26	ug/L	104.4	75 - 125				10/16/12
1,1-Dichloroethane	75-34-3		25	ug/L	101	75 - 125				10/16/12
Ethylbenzene	100-41-4		26	ug/L	104.5	75 - 125				10/16/12
Styrene	100-42-5		28	ug/L	110.6	75 - 125				10/16/12
trans-1,3-Dichloropropene	10061-02-6		26	ug/L	104.6	75 - 125				10/16/12
1,2-Dichloroethane	107-06-2		27	ug/L	106.5	75 - 125				10/16/12
1,1,1-Trichloroethane	71-55-6		26	ug/L	104.9	75 - 125				10/16/12
Dibromochloromethane	124-48-1		28	ug/L	110	75 - 125				10/16/12
Carbon disulfide	75-15-0		25	ug/L	98.6	75 - 125				10/16/12
Bromoform	75-25-2		30	ug/L	121.3	75 - 125				10/16/12
Bromodichloromethane	75-27-4		27	ug/L	107.4	75 - 125				10/16/12
1,2-Dichloropropane	78-87-5		27	ug/L	106.6	75 - 125				10/16/12
1,1,2-Trichloroethane	79-00-5		27	ug/L	109.8	75 - 125				10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
1,1,2,2-Tetrachloroethane	79-34-5		28	ug/L	111.6	75 - 125				10/16/12
trans-1,2-Dichloroethene	156-60-5		26	ug/L	104.4	75 - 125				10/16/12
cis-1,2-Dichloroethene	156-59-2		25	ug/L	98.5	75 - 125				10/16/12
MSD			**QC Sample #82692**							
			Original 121223013					**Paired**	**82691**	
1,1-Dichloroethene	75-35-4		25	ug/L	101.7	75 - 125	1.40	20		10/16/12
Trichloroethene	79-01-6		24	ug/L	97	75 - 125	2.00	20		10/16/12
Benzene	71-43-2		26	ug/L	102.2	75 - 125	3.00	20		10/16/12
Toluene	108-88-3		25	ug/L	99.2	75 - 125	2.50	20		10/16/12
Chlorobenzene	108-90-7		25	ug/L	100.5	75 - 125	3.80	20		10/16/12
1,1-Dichloroethane	75-34-3		25	ug/L	100.5	75 - 125	0.60	20		10/16/12
Ethylbenzene	100-41-4		26	ug/L	102.4	75 - 125	2.00	20		10/16/12
Styrene	100-42-5		27	ug/L	107.1	75 - 125	3.20	20		10/16/12
trans-1,3-Dichloropropene	10061-02-6		25	ug/L	100.4	75 - 125	4.10	20		10/16/12
1,2-Dichloroethane	107-06-2		25	ug/L	101.8	75 - 125	4.40	20		10/16/12
1,1,1-Trichloroethane	71-55-6		26	ug/L	103	75 - 125	1.80	20		10/16/12
Dibromochloromethane	124-48-1		26	ug/L	105.3	75 - 125	4.40	20		10/16/12
Carbon disulfide	75-15-0		25	ug/L	98.6	75 - 125	0.00	20		10/16/12
Bromoform	75-25-2		28	ug/L	113.4	75 - 125	6.80	20		10/16/12
Bromodichloromethane	75-27-4		26	ug/L	104.5	75 - 125	2.80	20		10/16/12
1,2-Dichloropropane	78-87-5		26	ug/L	103.3	75 - 125	3.10	20		10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
1,1,2-Trichloroethane	79-00-5		26	ug/L	104.2	75 - 125	5.20	20		10/16/12
1,1,2,2-Tetrachloroethane	79-34-5		26	ug/L	103.4	75 - 125	7.60	20		10/16/12
trans-1,2-Dichloroethene	156-60-5		24	ug/L	97.2	75 - 125	7.10	20		10/16/12
cis-1,2-Dichloroethene	156-59-2		25	ug/L	98.6	75 - 125	0.10	20		10/16/12

* - QC result out of range n/a - Not Applicable

Quality Control Report

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Analytical Batch 208482 (QC Batch: 208476) **Test** Cyanide (W) by Midi/Spectrophotometer
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82749**							
Cyanide	57-12-5		<4.0	ug/L					U	10/10/12
LCS			**QC Sample #82752**							
Cyanide	57-12-5		52.4	ug/L	104.8	85 - 115				10/10/12
MS			**QC Sample #82753**							
			Original 121203010							
Cyanide	57-12-5		40.4	ug/L	100.9	75 - 125				10/10/12
MSD			**QC Sample #82754**							
			Original 121203010					**Paired**	**82753**	
Cyanide	57-12-5		40.8	ug/L	102	75 - 125	1.10	20		10/10/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analytical Batch 208492 (QC Batch: 208454) **Test** ICP-2008 MS All possible metal
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82681**						
Aluminum	7429-90-5		<5.0	ug/L				U	10/15/12
Manganese	7439-96-5		<0.10	ug/L				U	10/15/12
Nickel	7440-02-0		<0.10	ug/L				U	10/15/12
Silver	7440-22-4		<0.050	ug/L				U	10/15/12
Antimony	7440-36-0		<0.30	ug/L				U	10/15/12
Barium	7440-39-3		0.218	ug/L				B	10/15/12
Beryllium	7440-41-7		<0.10	ug/L				U	10/15/12
Cadmium	7440-43-9		<0.050	ug/L				U	10/15/12
Chromium	7440-47-3		<0.10	ug/L				U	10/15/12
Cobalt	7440-48-4		<0.050	ug/L				U	10/15/12
Copper	7440-50-8		<0.10	ug/L				U	10/15/12
Vanadium	7440-62-2		0.454	ug/L				B	10/15/12
Zinc	7440-66-6		<1.0	ug/L				U	10/15/12
Lead	7439-92-1		0.0553	ug/L				B	10/15/12
Mercury	7439-97-6		<0.050	ug/L				U	10/15/12
Molybdenum	7439-98-7		0.0581	ug/L				B	10/15/12
Strontium	7440-24-6		0.619	ug/L				B	10/15/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Thallium	7440-28-0		<0.050	ug/L					U	10/15/12
Tin	7440-31-5		<0.050	ug/L					U	10/15/12
Arsenic	7440-38-2		<0.20	ug/L					U	10/15/12
Selenium	7782-49-2		<1.0	ug/L					U	10/15/12
LCS			**QC Sample #82682**							
Aluminum	7429-90-5		427	ug/L	106.7	85 - 115				10/15/12
Manganese	7439-96-5		41.6	ug/L	104	85 - 115				10/15/12
Nickel	7440-02-0		40.9	ug/L	102.2	85 - 115				10/15/12
Silver	7440-22-4		40.6	ug/L	101.6	85 - 115				10/15/12
Antimony	7440-36-0		39.0	ug/L	97.6	85 - 115				10/15/12
Barium	7440-39-3		41.6	ug/L	103.9	85 - 115				10/15/12
Beryllium	7440-41-7		39.4	ug/L	98.5	85 - 115				10/15/12
Cadmium	7440-43-9		39.4	ug/L	98.6	85 - 115				10/15/12
Chromium	7440-47-3		41.0	ug/L	102.6	85 - 115				10/15/12
Cobalt	7440-48-4		41.0	ug/L	102.6	85 - 115				10/15/12
Copper	7440-50-8		40.6	ug/L	101.5	85 - 115				10/15/12
Vanadium	7440-62-2		42.0	ug/L	105	85 - 115				10/15/12
Zinc	7440-66-6		37.2	ug/L	93	85 - 115				10/15/12
Lead	7439-92-1		42.8	ug/L	107	85 - 115				10/15/12
Mercury	7439-97-6		1.79	ug/L	89.4	85 - 115				10/15/12
Molybdenum	7439-98-7		40.8	ug/L	101.9	85 - 115				10/15/12
Strontium	7440-24-6		409	ug/L	102.2	85 - 115				10/15/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Thallium	7440-28-0		42.3	ug/L	105.7	85 - 115				10/15/12
Tin	7440-31-5		40.9	ug/L	102.3	85 - 115				10/15/12
Arsenic	7440-38-2		38.5	ug/L	96.2	85 - 115				10/15/12
Selenium	7782-49-2		35.7	ug/L	89.2	85 - 115				10/15/12
MS			**QC Sample #82683**							
			Original 121283007							
Aluminum	7429-90-5		402	ug/L	100.6	70 - 130				10/15/12
Manganese	7439-96-5		39.0	ug/L	97.6	70 - 130				10/15/12
Nickel	7440-02-0		36.2	ug/L	90.5	70 - 130				10/15/12
Silver	7440-22-4		38.1	ug/L	95.2	70 - 130				10/15/12
Antimony	7440-36-0		39.7	ug/L	99.2	70 - 130				10/15/12
Barium	7440-39-3		40.2	ug/L	100.6	70 - 130				10/15/12
Beryllium	7440-41-7		38.7	ug/L	96.6	70 - 130				10/15/12
Cadmium	7440-43-9		38.7	ug/L	96.8	70 - 130				10/15/12
Chromium	7440-47-3		38.9	ug/L	97.4	70 - 130				10/15/12
Cobalt	7440-48-4		37.7	ug/L	94.2	70 - 130				10/15/12
Copper	7440-50-8		35.5	ug/L	88.8	70 - 130				10/15/12
Vanadium	7440-62-2		40.0	ug/L	100	70 - 130				10/15/12
Zinc	7440-66-6		33.0	ug/L	82.4	70 - 130				10/15/12
Lead	7439-92-1		43.0	ug/L	107.5	70 - 130				10/15/12
Mercury	7439-97-6		1.93	ug/L	96.3	70 - 130				10/15/12
Molybdenum	7439-98-7		42.0	ug/L	105	70 - 130				10/15/12
Strontium	7440-24-6		406	ug/L	101.6	70 - 130				10/15/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Thallium	7440-28-0		43.6	ug/L	109	70 - 130				10/15/12
Tin	7440-31-5		40.7	ug/L	101.7	70 - 130				10/15/12
Arsenic	7440-38-2		38.9	ug/L	97.3	70 - 130				10/15/12
Selenium	7782-49-2		36.8	ug/L	92.1	70 - 130				10/15/12
MSD			**QC Sample #82684**							
			Original 121283007					**Paired**	**82683**	
Aluminum	7429-90-5		410	ug/L	102.4	70 - 130	1.70	20		10/15/12
Manganese	7439-96-5		39.8	ug/L	99.5	70 - 130	1.90	20		10/15/12
Nickel	7440-02-0		37.1	ug/L	92.9	70 - 130	2.60	20		10/15/12
Silver	7440-22-4		38.9	ug/L	97.2	70 - 130	2.10	20		10/15/12
Antimony	7440-36-0		40.7	ug/L	101.8	70 - 130	2.60	20		10/15/12
Barium	7440-39-3		42.2	ug/L	105.6	70 - 130	2.00	20		10/15/12
Beryllium	7440-41-7		38.4	ug/L	95.9	70 - 130	0.80	20		10/15/12
Cadmium	7440-43-9		40.2	ug/L	100.5	70 - 130	3.70	20		10/15/12
Chromium	7440-47-3		39.9	ug/L	99.8	70 - 130	1.90	20		10/15/12
Cobalt	7440-48-4		38.4	ug/L	96.1	70 - 130	2.00	20		10/15/12
Copper	7440-50-8		36.4	ug/L	91	70 - 130	2.40	20		10/15/12
Vanadium	7440-62-2		41.2	ug/L	103	70 - 130	1.90	20		10/15/12
Zinc	7440-66-6		34.2	ug/L	85.4	70 - 130	3.30	20		10/15/12
Lead	7439-92-1		44.4	ug/L	111.1	70 - 130	3.30	20		10/15/12
Mercury	7439-97-6		2.01	ug/L	100.3	70 - 130	4.10	20		10/15/12
Molybdenum	7439-98-7		43.0	ug/L	107.5	70 - 130	2.10	20		10/15/12
Strontium	7440-24-6		418	ug/L	104.6	70 - 130	1.80	20		10/15/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Thallium	7440-28-0		45.0	ug/L	112.6	70 - 130	3.20	20		10/15/12
Tin	7440-31-5		41.7	ug/L	104.2	70 - 130	2.40	20		10/15/12
Arsenic	7440-38-2		39.5	ug/L	98.8	70 - 130	1.50	20		10/15/12
Selenium	7782-49-2		36.8	ug/L	92	70 - 130	0.10	20		10/15/12

* - QC result out of range n/a - Not Applicable

Quality Control Report

Attention Scot Fitzgerald
Department Wet Chemistry

Group # WSCF121232

Analytical Batch 208754 (QC Batch: 208754) **Test** Total Alkalinity as mg/L CaCO3 (Water)
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
LCS			**QC Sample #82975**							
Total Alkalinity as CaCO3	ALKALINITY		97	mg/L	97	80 - 120				10/15/12
DUP			**QC Sample #82976**							
			Original 121239002							
Total Alkalinity as CaCO3	ALKALINITY		120	mg/L			0.00	20		10/15/12
LCS			**QC Sample #82977**							
Total Alkalinity as CaCO3	ALKALINITY		97	mg/L	97.5	80 - 120				10/15/12
LCS			**QC Sample #82978**							
Total Alkalinity as CaCO3	ALKALINITY		97	mg/L	97.2	80 - 120				10/15/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analytical Batch 208757 (QC Batch: 208756) **Test** Gasoline Range (W)
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit		RQ	Analyzed
BLANK			**QC Sample #82983**								
Gasoline	TPHGASOLI		<50	ug/L						U	10/16/12
LCS			**QC Sample #82984**								
Gasoline	TPHGASOLI		2300	ug/L	93.6	80 - 120					10/16/12
MS			**QC Sample #82985**								
			Original 121223013								
Gasoline	TPHGASOLI		1900	ug/L	78	75 - 125					10/16/12
MSD			**QC Sample #82986**								
			Original 121223013						**Paired**	**82985**	
Gasoline	TPHGASOLI		2100	ug/L	85.1	75 - 125	8.70	20			10/16/12
DUP			**QC Sample #82987**								
			Original 121223013								
Gasoline	TPHGASOLI		<50	ug/L			0.00	20		U	10/16/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analytical Batch 208850 (QC Batch: 208488) **Test** SW-846 8270D Semivolatiles
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82826**						
4-Nitrophenol	100-02-7		<1	ug/L				U	10/16/12
Phenol	108-95-2		<1	ug/L				U	10/16/12
1,2,4-Trichlorobenzene	120-82-1		<1	ug/L				U	10/16/12
2,4-Dinitrotoluene	121-14-2		<1	ug/L				U	10/16/12
1,4-Dichlorobenzene	106-46-7		<1	ug/L				U	10/16/12
Pyrene	129-00-0		<1	ug/L				U	10/16/12
4-Chloro-3-methylphenol	59-50-7		<1	ug/L				U	10/16/12
n-Nitroso-di-n-propylamine	621-64-7		<1	ug/L				U	10/16/12
Acenaphthene	83-32-9		<1	ug/L				U	10/16/12
Pentachlorophenol	87-86-5		<1	ug/L				U	10/16/12
2-Chlorophenol	95-57-8		<1	ug/L				U	10/16/12
4-Nitroaniline	100-01-6		<1	ug/L				U	10/16/12
4-Bromophenyl-phenylether	101-55-3		<1	ug/L				U	10/16/12
2,4-Dimethylphenol	105-67-9		<2	ug/L				U	10/16/12
4-Chloroaniline	106-47-8		<1	ug/L				U	10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Bis(1-Chloro-2-propyl)ether	108-60-1		<1	ug/L					U	10/16/12
Bis-(2-Chloroethyl) ether	111-44-4		<1	ug/L					U	10/16/12
Bis-(2-Chloroethoxy)methane	111-91-1		<1	ug/L					U	10/16/12
Bis-(2-Ethylhexyl)phthalate	117-81-7		<1	ug/L					U	10/16/12
Di-n-octylphthalate	117-84-0		<1	ug/L					U	10/16/12
Hexachlorobenzene	118-74-1		<1	ug/L					U	10/16/12
Anthracene	120-12-7		<1	ug/L					U	10/16/12
2,4-Dichlorophenol	120-83-2		<1	ug/L					U	10/16/12
Dimethylphthalate	131-11-3		<1	ug/L					U	10/16/12
Dibenzofuran	132-64-9		<1	ug/L					U	10/16/12
Benzo(g,h,i)perylene	191-24-2		<1	ug/L					U	10/16/12
Indeno(1,2,3-cd)pyrene	193-39-5		<1	ug/L					U	10/16/12
Benzo(b)fluoranthene	205-99-2		<1	ug/L					U	10/16/12
Fluoranthene	206-44-0		<1	ug/L					U	10/16/12
Benzo(k)fluoranthene	207-08-9		<1	ug/L					U	10/16/12
Acenaphthylene	208-96-8		<1	ug/L					U	10/16/12
Chrysene	218-01-9		<1	ug/L					U	10/16/12
Benzo(a)pyrene	50-32-8		<1	ug/L					U	10/16/12
2,4-Dinitrophenol	51-28-5		<1	ug/L					U	10/16/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov Limits	RPD	RPD Limit	RQ	Analyzed
Dibenzo(a,h)anthracene	53-70-3		<1	ug/L				U	10/16/12
4,6-Dinitro-2-methylphenol	534-52-1		<1	ug/L				U	10/16/12
1,3-Dichlorobenzene	541-73-1		<1	ug/L				U	10/16/12
Benzo(a)anthracene	56-55-3		<1	ug/L				U	10/16/12
2,6-Dinitrotoluene	606-20-2		<1	ug/L				U	10/16/12
4-Chlorophenyl-phenylether	7005-72-3		<1	ug/L				U	10/16/12
Hexachlorocyclopentadiene	77-47-4		<1	ug/L				U	10/16/12
Isophorone	78-59-1		<1	ug/L				U	10/16/12
Diethyl phthalate	84-66-2		<1	ug/L				U	10/16/12
Di-n-butylphthalate	84-74-2		<1	ug/L				U	10/16/12
Phenanthrene	85-01-8		<1	ug/L				U	10/16/12
Butylbenzylphthalate	85-68-7		<1	ug/L				U	10/16/12
Fluorene	86-73-7		<1	ug/L				U	10/16/12
Carbazole	86-74-8		<1	ug/L				U	10/16/12
Hexachlorobutadiene	87-68-3		<1	ug/L				U	10/16/12
2-Nitroaniline	88-74-4		<1	ug/L				U	10/16/12
2-Nitrophenol	88-75-5		<1	ug/L				U	10/16/12
Naphthalene	91-20-3		<1	ug/L				U	10/16/12
2-Methylnaphthalene	91-57-6		<1	ug/L				U	10/16/12
2-Chloronaphthalene	91-58-7		<1	ug/L				U	10/16/12
3,3-Dichlorobenzidine	91-94-1		<1	ug/L				U	10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
2-Methylphenol	95-48-7		<1	ug/L					U	10/16/12
1,2-Dichlorobenzene	95-50-1		<1	ug/L					U	10/16/12
2,4,5-Trichlorophenol	95-95-4		<1	ug/L					U	10/16/12
Nitrobenzene	98-95-3		<1	ug/L					U	10/16/12
3-Nitroaniline	99-09-2		<1	ug/L					U	10/16/12
3 & 4 Methylphenol, Total	65794-96-9		<1	ug/L					U	10/16/12
Hexachloroethane	67-72-1		<1	ug/L					U	10/16/12
2,4,6-Trichlorophenol	88-06-2		<1	ug/L					U	10/16/12
Benzyl alcohol	100-51-6		<1	ug/L					U	10/16/12
Tributyl phosphate	126-73-8		<1	ug/L					U	10/16/12
2-Naphthylamine	91-59-8		<2	ug/L					U	10/16/12
Pyridine	110-86-1		<1	ug/L					U	10/16/12
n-Nitrosopiperidine	100-75-4		<1	ug/L					U	10/16/12
n-Nitrosomethylethylamine	10595-95-6		<1	ug/L					U	10/16/12
p-Phenylenediamine	106-50-3		<1	ug/L					U	10/16/12
2-Picoline	109-06-8		<1	ug/L					U	10/16/12
3,3-Dimethylbenzidine	119-93-7		<4	ug/L					U	10/16/12
Isosafrole	120-58-1		<1	ug/L					U	10/16/12
Phentermine	122-09-8		<5	ug/L					U	10/16/12
1,4-Dioxane	123-91-1		<1	ug/L					U	10/16/12
1,4-Naphthoquinone	130-15-4		<1	ug/L					U	10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov Limits	RPD	RPD Limit	RQ	Analyzed
1-Naphthylamine	134-32-7		<2	ug/L				U	10/16/12
Aramite	140-57-8		<1	ug/L				U	10/16/12
Kepone	143-50-0		<1	ug/L				U	10/16/12
Hexachloropropene	1888-71-7		<1	ug/L				U	10/16/12
Diallate	2303-16-4		<1	ug/L				U	10/16/12
Pronamide	23950-58-5		<1	ug/L				U	10/16/12
Isodrin	465-73-6		<1	ug/L				U	10/16/12
Chlorobenzilate	510-15-6		<1	ug/L				U	10/16/12
2-Acetylaminofluorene	53-96-3		<1	ug/L				U	10/16/12
n-Nitrosodiethylamine	55-18-5		<1	ug/L				U	10/16/12
3-Methylcholanthrene	56-49-5		<1	ug/L				U	10/16/12
4-Nitroquinoline-1-oxide	56-57-5		<1	ug/L				U	10/16/12
7,12-Dimethylbenz(a)anthracene	57-97-6		<1	ug/L				U	10/16/12
2,3,4,6-Tetrachlorophenol	58-90-2		<1	ug/L				U	10/16/12
n-Nitrosomorpholine	59-89-2		<1	ug/L				U	10/16/12
Pentachlorobenzene	608-93-5		<1	ug/L				U	10/16/12
Phenacetin	62-44-2		<1	ug/L				U	10/16/12
Ethyl methanesulfonate	62-50-0		<1	ug/L				U	10/16/12
Aniline	62-53-3		<1	ug/L				U	10/16/12
n-Nitrosodimethylamine	62-75-9		<1	ug/L				U	10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Methyl methanesulfonate	66-27-3		<1	ug/L					U	10/16/12
Pentachloroethane	76-01-7		<1	ug/L					U	10/16/12
Pentachloronitrobenzene	82-68-8		<2	ug/L					U	10/16/12
2,6-Dichlorophenol	87-65-0		<1	ug/L					U	10/16/12
Dinoseb(..dinitromethylphenol)	88-85-7		<1	ug/L					U	10/16/12
4-Aminobiphenyl	92-67-1		<2	ug/L					U	10/16/12
n-Nitrosodibutylamine	924-16-3		<1	ug/L					U	10/16/12
n-Nitrosopyrolidine	930-55-2		<1	ug/L					U	10/16/12
Safrole	94-59-7		<1	ug/L					U	10/16/12
o-Toluidine	95-53-4		<1	ug/L					U	10/16/12
1,2,4,5-Tetrachlorobenzene	95-94-3		<1	ug/L					U	10/16/12
Acetophenone	98-86-2		<1	ug/L					U	10/16/12
1,3,5-Trinitrobenzene	99-35-4		<1	ug/L					U	10/16/12
2-Methyl-5-nitroaniline	99-55-8		<1	ug/L					U	10/16/12
1,3-Dinitrobenzene	99-65-0		<1	ug/L					U	10/16/12
O,O,O-Triethylthiophosphate	126-68-1		<1	ug/L					U	10/16/12
Parathion	56-38-2		<1	ug/L					U	10/16/12
Dimethylaminoazobenzene	60-11-7		<1	ug/L					U	10/16/12
Dimethoate	60-51-5		<1	ug/L					U	10/16/12
Thionazin	297-97-2		<1	ug/L					U	10/16/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Methyl parathion	298-00-0		<1	ug/L					U	10/16/12
Phorate	298-02-2		<1	ug/L					U	10/16/12
Disulfoton	298-04-4		<1	ug/L					U	10/16/12
Sulfotep	3689-24-5		<1	ug/L					U	10/16/12
Famfur	52-85-7		<5	ug/L					U	10/16/12
N-Nitrosodiphenylamin/ Diphenyl	DPA+NNDPA		<1	ug/L					U	10/16/12
Methapyrilene	91-80-5		<1	ug/L					U	10/16/12
LCS			**QC Sample #82827**							
4-Nitrophenol	100-02-7		14	ug/L	45.3	5 - 88				10/16/12
1,2,4-Trichlorobenzene	120-82-1		22	ug/L	73.4	50 - 105				10/16/12
Phenol	108-95-2		15	ug/L	48.7	18 - 89				10/16/12
1,4-Dichlorobenzene	106-46-7		15	ug/L	76.4	47 - 115				10/16/12
2,4-Dinitrotoluene	121-14-2		24	ug/L	81.3	59 - 110				10/16/12
Pyrene	129-00-0		26	ug/L	87.6	64 - 116				10/16/12
4-Chloro-3-methylphenol	59-50-7		25	ug/L	84.4	62 - 109				10/16/12
n-Nitroso-di-n-propylamine	621-64-7		24	ug/L	81.2	61 - 110				10/16/12
Acenaphthene	83-32-9		24	ug/L	78.6	59 - 113				10/16/12
Pentachlorophenol	87-86-5		22	ug/L	72.2	17 - 125				10/16/12
2-Chlorophenol	95-57-8		24	ug/L	78.4	55 - 109				10/16/12
1,4-Dioxane	123-91-1		20	ug/L	66.4	42 - 99				10/16/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
n-Nitrosodimethylamine	62-75-9		21	ug/L	71.3	40 - 103				10/16/12
Benzyl alcohol	100-51-6		25	ug/L	82.3	58 - 108				10/16/12
2-Methylphenol	95-48-7		24	ug/L	79.4	59 - 107				10/16/12
Hexachloroethane	67-72-1		20	ug/L	66	43 - 105				10/16/12
2-Nitrophenol	88-75-5		23	ug/L	77.6	48 - 113				10/16/12
2,4-Dimethylphenol	105-67-9		25	ug/L	84.6	58 - 113				10/16/12
2,4-Dichlorophenol	120-83-2		23	ug/L	78.2	52 - 110				10/16/12
Anthracene	120-12-7		26	ug/L	85.8	67 - 113				10/16/12
Naphthalene	91-20-3		23	ug/L	75.3	55 - 110				10/16/12
2-Nitroaniline	88-74-4		26	ug/L	87.1	57 - 114				10/16/12
Dibenzofuran	132-64-9		25	ug/L	82.4	61 - 113				10/16/12
Fluorene	86-73-7		25	ug/L	83.1	64 - 115				10/16/12
Tributyl phosphate	126-73-8		26	ug/L	87	65 - 108				10/16/12
Hexachlorobenzene	118-74-1		25	ug/L	84.2	60 - 117				10/16/12
Dimethoate	60-51-5		13	ug/L	86.9	64 - 108				10/16/12
Carbazole	86-74-8		27	ug/L	91.5	35 - 129				10/16/12
Di-n-butylphthalate	84-74-2		27	ug/L	88.3	70 - 116				10/16/12
3,3-Dichlorobenzidine	91-94-1		18	ug/L	58.5	16 - 117				10/16/12
Bis-(2-Ethylhexyl)phthalate	117-81-7		28	ug/L	93	64 - 133				10/16/12
Di-n-octylphthalate	117-84-0		25	ug/L	83	57 - 134				10/16/12
Benzo(a)pyrene	50-32-8		26	ug/L	88.1	63 - 115				10/16/12
2-Picoline	109-06-8		24	ug/L	80	59 - 102				10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Bis(1-Chloro-2-propyl)ether	108-60-1		23	ug/L	76	58 - 111				10/16/12
4-Chloroaniline	106-47-8		26	ug/L	85.7	43 - 125				10/16/12
MS			**QC Sample #82828**							
			Original 121239001							
4-Nitrophenol	100-02-7		7.7	ug/L	27.3	15 - 57				10/16/12
1,2,4-Trichlorobenzene	120-82-1		19	ug/L	66.3	51 - 104				10/16/12
Phenol	108-95-2		9.9	ug/L	34.9	24 - 65				10/16/12
1,4-Dichlorobenzene	106-46-7		13	ug/L	70	52 - 114				10/16/12
2,4-Dinitrotoluene	121-14-2		21	ug/L	73.1	57 - 112				10/16/12
Pyrene	129-00-0		23	ug/L	79.9	58 - 119				10/16/12
4-Chloro-3-methylphenol	59-50-7		21	ug/L	75.2	56 - 115				10/16/12
n-Nitroso-di-n-propylamine	621-64-7		21	ug/L	72.5	60 - 112				10/16/12
Acenaphthene	83-32-9		20	ug/L	69.9	60 - 113				10/16/12
Pentachlorophenol	87-86-5		13	ug/L	45	32 - 127				10/16/12
2-Chlorophenol	95-57-8		20	ug/L	69.3	52 - 113				10/16/12
1,4-Dioxane	123-91-1		16	ug/L	56.8	39 - 93				10/16/12
n-Nitrosodimethylamine	62-75-9		17	ug/L	61	41 - 92				10/16/12
Benzyl alcohol	100-51-6		21	ug/L	73.4	56 - 107				10/16/12
2-Methylphenol	95-48-7		19	ug/L	67.1	46 - 114				10/16/12
Hexachloroethane	67-72-1		17	ug/L	58.9	48 - 102				10/16/12
2-Nitrophenol	88-75-5		19	ug/L	67.3	51 - 114				10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit		RQ	Analyzed
2,4-Dimethylphenol	105-67-9		21	ug/L	74.5	46 - 124					10/16/12
2,4-Dichlorophenol	120-83-2		19	ug/L	68.8	50 - 114					10/16/12
Anthracene	120-12-7		22	ug/L	77	64 - 116					10/16/12
Naphthalene	91-20-3		19	ug/L	67	57 - 110					10/16/12
2-Nitroaniline	88-74-4		22	ug/L	76.1	60 - 114					10/16/12
Dibenzofuran	132-64-9		21	ug/L	73.9	61 - 114					10/16/12
Fluorene	86-73-7		21	ug/L	73.9	63 - 116					10/16/12
Tributyl phosphate	126-73-8		23	ug/L	79.7	59 - 113					10/16/12
Hexachlorobenzene	118-74-1		22	ug/L	76.3	58 - 119					10/16/12
Dimethoate	60-51-5		11	ug/L	76.3	53 - 119					10/16/12
Carbazole	86-74-8		22	ug/L	79.3	41 - 122					10/16/12
Di-n-butylphthalate	84-74-2		23	ug/L	80.2	67 - 118					10/16/12
3,3-Dichlorobenzidine	91-94-1		18	ug/L	62.9	16 - 121					10/16/12
Bis-(2-Ethylhexyl)phthalate	117-81-7		24	ug/L	85.1	64 - 134					10/16/12
Di-n-octylphthalate	117-84-0		22	ug/L	77.6	40 - 143					10/16/12
Benzo(a)pyrene	50-32-8		22	ug/L	79.1	61 - 117					10/16/12
2-Picoline	109-06-8		20	ug/L	70.9	50 - 104					10/16/12
Bis(1-Chloro-2-propyl)ether	108-60-1		19	ug/L	67.5	58 - 112					10/16/12
4-Chloroaniline	106-47-8		25	ug/L	87	43 - 118					10/16/12
MSD			**QC Sample #82829**								
			Original 121239001						**Paired**	**82828**	
4-Nitrophenol	100-02-7		11	ug/L	40.6	15 - 57	39.00	20	*	X	10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
1,2,4-Trichlorobenzene	120-82-1		20	ug/L	69.3	51 - 104	4.50	20		10/16/12
Phenol	108-95-2		11	ug/L	38.4	24 - 65	9.50	20		10/16/12
1,4-Dichlorobenzene	106-46-7		13	ug/L	71.3	52 - 114	1.90	20		10/16/12
2,4-Dinitrotoluene	121-14-2		22	ug/L	79.4	57 - 112	8.30	20		10/16/12
Pyrene	129-00-0		22	ug/L	76.3	58 - 119	4.60	20		10/16/12
4-Chloro-3-methylphenol	59-50-7		23	ug/L	80.6	56 - 115	6.90	20		10/16/12
n-Nitroso-di-n-propylamine	621-64-7		22	ug/L	76.8	60 - 112	5.80	20		10/16/12
Acenaphthene	83-32-9		21	ug/L	74.1	60 - 113	5.80	20		10/16/12
Pentachlorophenol	87-86-5		19	ug/L	67.7	32 - 127	40.30	20	* X	10/16/12
2-Chlorophenol	95-57-8		21	ug/L	73.4	52 - 113	5.70	20		10/16/12
1,4-Dioxane	123-91-1		18	ug/L	63.2	39 - 93	10.80	20		10/16/12
n-Nitrosodimethylamine	62-75-9		19	ug/L	68.3	41 - 92	11.20	20		10/16/12
Benzyl alcohol	100-51-6		23	ug/L	80.5	56 - 107	9.30	20		10/16/12
2-Methylphenol	95-48-7		21	ug/L	72.9	46 - 114	8.20	20		10/16/12
Hexachloroethane	67-72-1		18	ug/L	61.9	48 - 102	5.00	20		10/16/12
2-Nitrophenol	88-75-5		20	ug/L	72.3	51 - 114	7.20	20		10/16/12
2,4-Dimethylphenol	105-67-9		22	ug/L	79.4	46 - 124	6.30	20		10/16/12
2,4-Dichlorophenol	120-83-2		21	ug/L	73.1	50 - 114	6.10	20		10/16/12
Anthracene	120-12-7		23	ug/L	80	64 - 116	3.80	20		10/16/12
Naphthalene	91-20-3		20	ug/L	70.9	57 - 110	5.60	20		10/16/12
2-Nitroaniline	88-74-4		23	ug/L	82.3	60 - 114	7.80	20		10/16/12

* - QC result out of range

n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Dibenzofuran	132-64-9		22	ug/L	77.8	61 - 114	5.20	20		10/16/12
Fluorene	86-73-7		22	ug/L	79.3	63 - 116	7.00	20		10/16/12
Tributyl phosphate	126-73-8		23	ug/L	81.6	59 - 113	2.40	20		10/16/12
Hexachlorobenzene	118-74-1		22	ug/L	78.4	58 - 119	2.70	20		10/16/12
Dimethoate	60-51-5		12	ug/L	83.5	53 - 119	9.00	20		10/16/12
Carbazole	86-74-8		25	ug/L	87.9	41 - 122	10.30	20		10/16/12
Di-n-butylphthalate	84-74-2		24	ug/L	83.3	67 - 118	3.90	20		10/16/12
3,3-Dichlorobenzidine	91-94-1		19	ug/L	66.7	16 - 121	6.00	20		10/16/12
Bis-(2-Ethylhexyl)phthalate	117-81-7		23	ug/L	81.6	64 - 134	4.20	20		10/16/12
Di-n-octylphthalate	117-84-0		21	ug/L	75	40 - 143	3.50	20		10/16/12
Benzo(a)pyrene	50-32-8		23	ug/L	82.9	61 - 117	4.70	20		10/16/12
2-Picoline	109-06-8		23	ug/L	82	50 - 104	14.50	20		10/16/12
Bis(1-Chloro-2-propyl)ether	108-60-1		20	ug/L	70.3	58 - 112	4.00	20		10/16/12
4-Chloroaniline	106-47-8		26	ug/L	92.7	43 - 118	6.30	20		10/16/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analytical Batch 208887 (QC Batch: 208650) **Test** PCBs by EPA SW-846 Method 8082
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
BLANK			**QC Sample #82953**							
Aroclor-1016	12674-11-2		<0.1	ug/L					U	10/17/12
Aroclor-1221	11104-28-2		<0.2	ug/L					U	10/17/12
Aroclor-1232	11141-16-5		<0.1	ug/L					U	10/17/12
Aroclor-1242	53469-21-9		<0.1	ug/L					U	10/17/12
Aroclor-1248	12672-29-6		<0.1	ug/L					U	10/17/12
Aroclor-1254	11097-69-1		<0.1	ug/L					U	10/17/12
Aroclor-1260	11096-82-5		<0.1	ug/L					U	10/17/12
LCS			**QC Sample #82954**							
Aroclor-1254	11097-69-1		1.7	ug/L	87.2	70 - 130				10/17/12
MS			**QC Sample #82955**							
			Original 121223013							
Aroclor-1254	11097-69-1		1.8	ug/L	93.6	60 - 130				10/17/12
MSD			**QC Sample #82956**							
			Original 121223013					**Paired 82955**		
Aroclor-1254	11097-69-1		1.8	ug/L	95.1	60 - 130	1.60	20		10/17/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Inorganic

Group # WSCF121232

Analytical Batch 209139 (QC Batch: 209136) **Test** Total Organic Halides
Associated Samples 121232001, 121232002, 121232005, 121232006, 121232007, 121232008, 121232009, 121232010

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit		RQ	Analyzed
BLANK			**QC Sample #83612**								
Total Organic Halides	59473-04-0		<5.0	ug/L						U	10/08/12
LCS			**QC Sample #83613**								
Total Organic Halides	59473-04-0		377	mg/L	94.3	80 - 120					10/08/12
MS			**QC Sample #83617**								
			Original 121232001								
Total Organic Halides	59473-04-0	<5.0	39.2	ug/L	98	75 - 125					10/08/12
MSD			**QC Sample #83618**								
			Original 121232001						**Paired**	**83617**	
Total Organic Halides	59473-04-0	<5.0	42.6	ug/L	106.4	75 - 125	8.20	20			10/08/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analytical Batch 208258 (QC Batch: 208245) **Test** Extractable Diesel and Petroleum
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
SAMPLE			**Sample #121232001**							
o-Terphenyl	84-15-1				107.3	70 - 130				10/05/12
SAMPLE			**Sample #121232002**							
o-Terphenyl	84-15-1				84.6	70 - 130				10/05/12
BLANK			**QC Sample #82350**							
o-Terphenyl	84-15-1				94.7	70 - 130				10/05/12
LCS			**QC Sample #82351**							
o-Terphenyl	84-15-1				102.3	70 - 130				10/05/12
MS			**QC Sample #82352**							
			Original 121223013							
o-Terphenyl	84-15-1				94.3	70 - 130				10/05/12
MSD			**QC Sample #82353**							
			Original 121223013					**Paired 82352**		
o-Terphenyl	84-15-1				96.6	70 - 130	n/a			10/05/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analytical Batch 208457 (QC Batch: 208456) **Test** SW-846 8260B Volatiles
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
SAMPLE			**Sample #121232001**							
1,2-Dichloroethane-d4	17060-07-0				105.2	75 - 125				10/16/12
Toluene-d8	2037-26-5				98.2	75 - 125				10/16/12
4-Bromofluorobenzene	460-00-4				100.5	75 - 125				10/16/12
SAMPLE			**Sample #121232002**							
1,2-Dichloroethane-d4	17060-07-0				101.1	75 - 125				10/16/12
Toluene-d8	2037-26-5				98.1	75 - 125				10/16/12
4-Bromofluorobenzene	460-00-4				100.4	75 - 125				10/16/12
BLANK			**QC Sample #82689**							
1,2-Dichloroethane-d4	17060-07-0				101.4	75 - 125				10/16/12
Toluene-d8	2037-26-5				98.7	75 - 125				10/16/12
4-Bromofluorobenzene	460-00-4				98.6	75 - 125				10/16/12
LCS			**QC Sample #82690**							
1,2-Dichloroethane-d4	17060-07-0				107.9	75 - 125				10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Toluene-d8	2037-26-5				96.4	75 - 125				10/16/12
4-Bromofluorobenzene	460-00-4				95	75 - 125				10/16/12
MS			**QC Sample #82691**							
			Original 121223013							
4-Bromofluorobenzene	460-00-4				95.3	75 - 125				10/16/12
1,2-Dichloroethane-d4	17060-07-0				104.4	75 - 125				10/16/12
Toluene-d8	2037-26-5				97	75 - 125				10/16/12
MSD			**QC Sample #82692**							
			Original 121223013					**Paired**	**82691**	
1,2-Dichloroethane-d4	17060-07-0				103.9	75 - 125	n/a			10/16/12
Toluene-d8	2037-26-5				97.5	75 - 125	n/a			10/16/12
4-Bromofluorobenzene	460-00-4				96.2	75 - 125	n/a			10/16/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analytical Batch 208757 (QC Batch: 208756) **Test** Gasoline Range (W)
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
SAMPLE			**Sample #121232001**							
4-Bromofluorobenzene	460-00-4				95.7	50 - 150				10/16/12
SAMPLE			**Sample #121232002**							
4-Bromofluorobenzene	460-00-4				93.9	50 - 150				10/16/12
BLANK			**QC Sample #82983**							
4-Bromofluorobenzene	460-00-4				99.1	50 - 150				10/16/12
LCS			**QC Sample #82984**							
4-Bromofluorobenzene	460-00-4				97.1	50 - 150				10/16/12
MS			**QC Sample #82985**							
			Original 121223013							
4-Bromofluorobenzene	460-00-4				96.7	50 - 150				10/16/12
MSD			**QC Sample #82986**							
			Original 121223013					**Paired 82985**		

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Volatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
4-Bromofluorobenzene	460-00-4				98.5	50 - 150	n/a			10/16/12
DUP			**QC Sample #82987**							
			Original 121223013							
4-Bromofluorobenzene	460-00-4				96.9	50 - 150	n/a			10/16/12

* - QC result out of range n/a - Not Applicable

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analytical Batch 208850 (QC Batch: 208488) **Test** SW-846 8270D Semivolatiles
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
SAMPLE			**Sample #121232001**							
2-Fluorophenol	367-12-4				50.7	44 - 135				10/16/12
Phenol-d5	4165-62-2				33.1	41 - 136			X	10/16/12
Nitrobenzene-d5	4165-60-0				73.1	53 - 129				10/16/12
2-Methylnaphthalene-d10	7297-45-2				76.3	50 - 140				10/16/12
2-Fluorobiphenyl	321-60-8				74.6	36 - 141				10/16/12
2,4,6-Tribromophenol	118-79-6				71.2	17 - 142				10/16/12
Fluoranthene-d10	93951-69-0				86.2	50 - 140				10/16/12
Terphenyl-d14	98904-43-9				76.8	61 - 142				10/16/12
SAMPLE			**Sample #121232002**							
2-Fluorophenol	367-12-4				50.2	44 - 135				10/16/12
Phenol-d5	4165-62-2				32.8	41 - 136			X	10/16/12
Nitrobenzene-d5	4165-60-0				73	53 - 129				10/16/12
2-Methylnaphthalene-d10	7297-45-2				73.5	50 - 140				10/16/12
2-Fluorobiphenyl	321-60-8				73.4	36 - 141				10/16/12
2,4,6-Tribromophenol	118-79-6				64.1	17 - 142				10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
Fluoranthene-d10	93951-69-0				77.9	50 - 140				10/16/12
Terphenyl-d14	98904-43-9				73.2	61 - 142				10/16/12
BLANK			**QC Sample #82826**							
2-Fluorophenol	367-12-4				56.1	44 - 135				10/16/12
Phenol-d5	4165-62-2				40.6	41 - 136			X	10/16/12
Nitrobenzene-d5	4165-60-0				73.6	53 - 129				10/16/12
2-Methylnaphthalene-d10	7297-45-2				74.2	50 - 140				10/16/12
2-Fluorobiphenyl	321-60-8				73.9	36 - 141				10/16/12
2,4,6-Tribromophenol	118-79-6				64.9	17 - 142				10/16/12
Fluoranthene-d10	93951-69-0				83.9	50 - 140				10/16/12
Terphenyl-d14	98904-43-9				76.9	61 - 142				10/16/12
LCS			**QC Sample #82827**							
2-Fluorophenol	367-12-4				64.3	44 - 135				10/16/12
Phenol-d5	4165-62-2				46.6	41 - 136				10/16/12
Nitrobenzene-d5	4165-60-0				81.8	53 - 129				10/16/12
2-Methylnaphthalene-d10	7297-45-2				81.7	50 - 140				10/16/12
2-Fluorobiphenyl	321-60-8				81.2	36 - 141				10/16/12
2,4,6-Tribromophenol	118-79-6				79.5	17 - 142				10/16/12
Fluoranthene-d10	93951-69-0				88.1	50 - 140				10/16/12
Terphenyl-d14	98904-43-9				90.7	61 - 142				10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
MS			**QC Sample #82828**							
			Original 121239001							
2-Fluorophenol	367-12-4				50.8	44 - 135				10/16/12
Phenol-d5	4165-62-2				33.2	41 - 136			X	10/16/12
Nitrobenzene-d5	4165-60-0				70.9	53 - 129				10/16/12
2-Methylnaphthalene-d10	7297-45-2				72.3	50 - 140				10/16/12
2-Fluorobiphenyl	321-60-8				72.6	36 - 141				10/16/12
2,4,6-Tribromophenol	118-79-6				70.6	17 - 142				10/16/12
Fluoranthene-d10	93951-69-0				77.1	50 - 140				10/16/12
Terphenyl-d14	98904-43-9				84.6	61 - 142				10/16/12
MSD			**QC Sample #82829**							
			Original 121239001					**Paired**	**82828**	
2-Fluorophenol	367-12-4				54.6	44 - 135	n/a			10/16/12
Phenol-d5	4165-62-2				37.3	41 - 136	n/a		X	10/16/12
Nitrobenzene-d5	4165-60-0				75.6	53 - 129	n/a			10/16/12
2-Methylnaphthalene-d10	7297-45-2				76.7	50 - 140	n/a			10/16/12
2-Fluorobiphenyl	321-60-8				75.4	36 - 141	n/a			10/16/12
2,4,6-Tribromophenol	118-79-6				76	17 - 142	n/a			10/16/12
Fluoranthene-d10	93951-69-0				85.5	50 - 140	n/a			10/16/12
Terphenyl-d14	98904-43-9				81.1	61 - 142	n/a			10/16/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analytical Batch 208887 (QC Batch: 208650) **Test** PCBs by EPA SW-846 Method 8082
Associated Samples 121232001, 121232002

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
SAMPLE			**Sample #121232001**							
Tetrachloro-m-xylene	877-09-8				78.5	60 - 140				10/17/12
Decachlorobiphenyl	2051-24-3				94.7	60 - 140				10/17/12
SAMPLE			**Sample #121232002**							
Tetrachloro-m-xylene	877-09-8				86.4	60 - 140				10/17/12
Decachlorobiphenyl	2051-24-3				94.7	60 - 140				10/17/12
BLANK			**QC Sample #82953**							
Tetrachloro-m-xylene	877-09-8				78.2	60 - 140				10/17/12
Decachlorobiphenyl	2051-24-3				91.9	60 - 140				10/17/12
LCS			**QC Sample #82954**							
Tetrachloro-m-xylene	877-09-8				75.5	60 - 140				10/17/12
Decachlorobiphenyl	2051-24-3				94	60 - 140				10/17/12
MS			**QC Sample #82955**							
			Original 121223013							
Tetrachloro-m-xylene	877-09-8				81.9	60 - 140				10/17/12
Decachlorobiphenyl	2051-24-3				95.8	60 - 140				10/17/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Quality Control Report

Attention Scot Fitzgerald
Department Organic, Semivolatiles

Group # WSCF121232

Analyte	CAS #	Original Found	QC Found	Units	% Recov	Limits	RPD	RPD Limit	RQ	Analyzed
MSD			**QC Sample #82956**							
			Original 121223013					**Paired 82955**		
Tetrachloro-m-xylene	877-09-8				81.2	60 - 140	n/a			10/17/12
Decachlorobiphenyl	2051-24-3				96.8	60 - 140	n/a			10/17/12

* - QC result out of range n/a - Not Applicable

REVISED121232 -

Attention: Scot Fitzgerald **Group #** WSCF121232

121232001	**B2M177**
Department	Organic, Semivolatiles
Analyte	Phenol-d5 - SW-846 8270D Semivolatiles [1] Surrogate recovery outside of established laboratory control limits.

121232002	**B2M9W6**
Department	Organic, Semivolatiles
Analyte	Phenol-d5 - SW-846 8270D Semivolatiles [1] Surrogate recovery outside of established laboratory control limits.

Attention: Scot Fitzgerald **Group #** WSCF121232

Quality Control Comments

Department Organic, Semivolatiles

82826 BLANK for HBN 208488 [ORGP/202

Analyte Phenol-d5 - SW-846 8270D Semivolatiles
[1] Surrogate recovery outside of established laboratory control limits.

82828 B2M0Y3(121239001MS)

Analyte Phenol-d5 - SW-846 8270D Semivolatiles
[1] Surrogate recovery outside of established laboratory control limits.

82829 B2M0Y3(121239001MSD)

Analyte 4-Nitrophenol - SW-846 8270D Semivolatiles
[1] Matrix Spike RPD outside established laboratory limits No flags assigned.

Analyte Pentachlorophenol - SW-846 8270D Semivolatiles
[1] Matrix Spike RPD outside established laboratory limits No flags assigned.

Analyte Phenol-d5 - SW-846 8270D Semivolatiles
[1] Surrogate recovery outside of established laboratory control limits.

ATTACHMENT4

SAMPLE RECEIPT

Consisting of 7 pages
Including cover page

Waste Sampling and Characterization Facility
P.O. Box 1970 S3-30, Richland WA 99352
Phone: (509) 373-7004/FAX: (509) 373-7134

ACKNOWLEDGEMENT OF SAMPLES RECEIVED

WSCF Laboratory
PO Box 650 S3-30
Richland, WA 99352

ATTN: Scot Fitzgerald

Customer Code: CHPRC
PO #: 401647
Work Order #: 121232
Profile #: W13-010-199
Proj. Mgr.:
Phone:

The following samples were received from you on 10/3/2012 2:30:00 PM. They have been scheduled for the tests listed beside each sample. If this information is incorrect, please contact your service representative. Thank you for using Waste Sampling and Characterization Facility.

Sample #	Sample ID	Matrix	Collected	Received
		Tests scheduled		
121232001	B2M177	WATER	10/3/2012 13:33	10/3/2012 14:30
		2008-W; 6010-W; 8260V-W; 8270SV-W; ALK-W; CN-W; COD-W; PCB-W; TOC-W; TOX-W; TPHDWA-W; TPHGWA-W		
121232002	B2M9W6	WATER	10/3/2012 12:03	10/3/2012 14:30
		2008-W; 6010-W; 8260V-W; 8270SV-W; ALK-W; CN-W; COD-W; PCB-W; TOC-W; TOX-W; TPHDWA-W; TPHGWA-W		
121232003	B2M179	WATER	10/3/2012 13:33	10/3/2012 14:30
		2008-W; 6010-W		
121232004	B2M9W5	WATER	10/3/2012 12:03	10/3/2012 14:30
		2008-W; 6010-W		
121232005	B2M239	WATER	10/3/2012 13:33	10/3/2012 14:30
		TOC-W; TOX-W		
121232006	B2M240	WATER	10/3/2012 13:33	10/3/2012 14:30
		TOC-W; TOX-W		
121232007	B2M241	WATER	10/3/2012 13:33	10/3/2012 14:30
		TOC-W; TOX-W		
121232008	B2M9W7	WATER	10/3/2012 12:03	10/3/2012 14:30
		TOC-W; TOX-W		
121232009	B2M9W8	WATER	10/3/2012 12:03	10/3/2012 14:30
		TOC-W; TOX-W		
121232010	B2M9W9	WATER	10/3/2012 12:03	10/3/2012 14:30
		TOC-W; TOX-W		

Waste Sampling and Characterization Facility
P.O. Box 1970 S3-30, Richland WA 99352
Phone: (509) 373-7004/FAX: (509) 373-7134

121232011	B2M8K7	WATER	10/3/2012 13:33	10/3/2012 14:30
		2008-W		
121232012	B2M8K8	WATER	10/3/2012 13:33	10/3/2012 14:30
		2008-W		

Test Acronym Description

Test Acronym	Description
2008-W	ICP-MS (W)
6010-W	ICP-AES (W)
8260V-W	Volatiles by 8260B (W)
8270SV-W	Semivolatiles by 8270D (W)
ALK-W	Total Alkalinity (W)
CN-W	Cyanide (Spectroscopy) (W)
COD-W	Chemical Oxygen Demand (W)
PCB-W	PCB (8082) (W)
TOC-W	Total Organic Carbon (W)
TOX-W	Total Organic Halides (W)
TPHDWA-W	TPHD-WA (W)
TPHGWA-W	TPHG-WA (Water)

CH2MHill Plateau Remediation Company	CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST	C.O.C. # W13-010-199 Page 1 of 2

Collector	D.J. Woehle CHPRC	Contact/Requester	Karen Waters-Husted	Telephone No.	376-4650
SAF No.	W13-010	Sampling Origin	Hanford Site	Purchase Order/Charge Code	300071ES20
Project Title	RCRA, OCTOBER 2012	Logbook No.	HNF-N-506 48168	Ice Chest No.	N/A
Shipped To (Lab)	Waste Sampling & Characterization	Method of Shipment	GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.	N/A
Protocol	RCRA	Priority: 31 Days	PRIORITY	Offsite Property No.	N/A

POSSIBLE SAMPLE HAZARDS/REMARKS

** ** Contains Radioactive Material at concentations that are not regulated for transportation per 49 CRF but are not releasable per DOE Order 5400 5 (1990/1993)

SPECIAL INSTRUCTIONS Hold Time — Total Activity Exemption: Yes ☑ No ☐

FY12 and FY13 samples cannot be in the same SDG.
Site Wide Generator Knowledge Information Form applies.
The CACN for all analytical work at WSCF is 401647.

LW. 10/3/12 121232

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B2M8K7	N	W	10/3/12	~~1333~~ 1033	1x500-mL G	200.8_HG - ICPMS	28 Days	HNO3 to pH <2
B2M240	N	W			1x1-L aGs*	9020_TOX: TOX (1)	28 Days	H2SO4 to pH <2/Cool~4C
B2M240	N	W			1x250-mL aG	9060_TOC: TOC (1)	28 Days	HCl or H2SO4 to pH <2/Cool~4C
B2M8K8	Y	W			1x500-mL G	200.8_HG - ICPMS	28 Days	HNO3 to pH <2
B2M241	N	W			1x1-L aGs*	9020_TOX: TOX (1)	28 Days	H2SO4 to pH <2/Cool~4C
B2M241	N	W			1x250-mL aG	9060_TOC: TOC (1)	28 Days	HCl or H2SO4 to pH <2/Cool~4C
B2M177	N	W			1x500-mL G/P	200.8_METALS_ICPMS: List-1 (26)	6 Months	HNO3 to pH <2
B2M177	N	W			1x250-mL G/P	2320_ALKALINITY: Alkalinity (1)	14 Days	Cool~4C
B2M177	N	W			1x500-mL G/P	410.4_COD: COD (1)	28 Days	H2SO4 to pH <2/Cool~4C
B2M177	N	W			1x250-mL P	4500E_CN: Cyanide (1)	14 Days	NaOH to pH >=12
B2M177	N	W			1x500-mL G/P	6010_METALS_ICP: List-3 (18)	6 Months	HNO3 to pH <2
B2M177	N	W			4x1-L aG	8082_PCB_GC: List-1 (7)	None	Cool~4C
B2M177	N	W			1x1-L aGs*	9020_TOX: TOX (1)	28 Days	H2SO4 to pH <2/Cool~4C
~~B2M177~~	~~N~~	~~W~~			~~1x250-mL G/P~~	~~9030_SULFIDE: Sulfide (1)~~ LW. 10/3/12	7 Days	~~ZnAc+NaOH to pH >9/Cool~4C~~

LW. 10/3/12

Relinquished By	Print	Sign	Date/Time	Received By	Print	Sign	Date/Time
D.J. Woehle CHPRC			OCT 03 2012 1430	C Johnson			OCT 03 2012 1430
Relinquished By			Date/Time	Received By			Date/Time
Relinquished By			Date/Time	Received By			Date/Time
Relinquished By			Date/Time	Received By			Date/Time

Matrix *

S = Soil	DS = Drum Solids		
SE = Sediment	DL = Drum Liquids		
SO = Solid	T = Tissue		
SL = Sludge	WI = Wipe		
W = Water	L = Liquid		
O = Oil	V = Vegetation		
A = Air	X = Other		

FINAL SAMPLE DISPOSITION	Disposal Method (e.g., Return to customer, per lab procedure, used in process)	Disposed By	Date/Time

PRINTED ON 9/18/2012

A-6004-842 (REV 2)

Chain of Custody

CH2MHill Plateau Remediation Company	CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST	C.O.C. # W13-010-199 Page 2 of 2

Collector	D.J. Woehle CHPRC	Contact/Requester	Karen Waters-Husted	Telephone No.	376-4650
SAF No.	W13-010	Sampling Origin	Hanford Site	Purchase Order/Charge Code	300071ES20
Project Title	RCRA, OCTOBER 2012	Logbook No.	HNF-N-506 48 / 68	Ice Chest No.	N/A
Shipped To (Lab)	Waste Sampling & Characterization	Method of Shipment	GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.	N/A
Protocol	RCRA	Priority:	31 Days PRIORITY	Offsite Property No.	N/A

POSSIBLE SAMPLE HAZARDS/REMARKS

** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CRF but are not releasable per DOE Order 5400.5 (1990/1993)

SPECIAL INSTRUCTIONS **Hold Time** Total Activity Exemption: Yes ☑ No ☐

FY12 and FY13 samples cannot be in the same SDG.
Site Wide Generator Knowledge Information Form applies.
The CACN for all analytical work at WSCF is 401647.

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B2M177	N	W	10/3/12	AW.10/3/12 10/1333	1x250-mL aG	9060_TOC: TOC (1)	28 Days	HCl or H2SO4 to pH <2/Cool~4C
B2M177	N	W			3x1-L aG	TPH-DieselKerosene Range - WTPH-D	14/40 Days	HCl to pH <2/Cool~4C
B2M177	N	W			4x40-mL aGs*	TPH-Gasoline Range - WTPH-G	14 Days	HCl to pH <2/Cool~4C
B2M177	N	W			3x40-mL aGs*	8260_VOA_GCMS_IX: COMMON; 8260_VOA_GCMS_IX: COMMON (Add-on)	14 Days	HCl or H2SO4 to pH <2/Cool~4C
B2M177	N	W			4x1-L aG	8270_SVOA_GCMS_IX: COMMON	7/40 Days	Cool~4C
B2M239 5	N	W			1x1-L aGs*	9020_TOX: TOX (1)	28 Days	H2SO4 to pH <2/Cool~4C
B2M239	N	W			1x250-mL aG	9060_TOC: TOC (1)	28 Days	HCl or H2SO4 to pH <2/Cool~4C
B2M179 3	Y	W			1x500-mL G/P	200.8_METALS_ICPMS: List-1 (26)	6 Months	HNO3 to pH <2
B2M179	Y	W			1x500-mL G/P	6010_METALS_ICP: List-3 (18)	6 Months	HNO3 to pH <2

Relinquished By	Date/Time	Received By	Date/Time
D.J. Woehle CHPRC (signed)	OCT 0 3 2012 1430	C. Dukhum (signed)	OCT 0 3 2012 1430
Relinquished By	Date/Time	Received By	Date/Time
Relinquished By	Date/Time	Received By	Date/Time
Relinquished By	Date/Time	Received By	Date/Time

Matrix *: S = Soil; SE = Sediment; SO = Solid; SL = Sludge; W = Water; O = Oil; A = Air; DS = Drum Solids; DL = Drum Liquids; T = Tissue; WI = Wipe; L = Liquid; V = Vegetation; X = Other

FINAL SAMPLE DISPOSITION Disposal Method (e.g., Return to customer, per lab procedure, used in process) Disposed By Date/Time

PRINTED ON 9/18/2012 A-6004-842 (REV 2)

Chain of Custody

CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. # **W13-010-303**

Page 1 of 2

Collector: D.J. Woehle CHPRC	Contact/Requester: Karen Waters-Husted	Telephone No.: 376-4650	
SAF No.: W13-010	Sampling Origin: Hanford Site	Purchase Order/Charge Code: 300071ES20	
Project Title: RCRA, OCTOBER 2012	Logbook No.: HNF-N-506 48/68	Ice Chest No.: N/A	
Shipped To (Lab): Waste Sampling & Characterization	Method of Shipment: GOVERNMENT VEHICLE	Bill of Lading/Air Bill No.: N/A	
Protocol: RCRA	Priority: 31 Days **PRIORITY**	Offsite Property No.: N/A	

POSSIBLE SAMPLE HAZARDS/REMARKS

** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CFR but are not releasable per DOE Order 5400.5 (1990/1993)

SPECIAL INSTRUCTIONS Hold Time

FY12 and FY13 samples cannot be in the same SDG.
Site Wide Generator Knowledge Information Form applies.
The CACN for all analytical work at WSCF is 401647.

Total Activity Exemption: Yes ☑ No ☐

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B2M9W5	Y	W	10/3/12	1203	1x500-mL G/P	200.8_METALS_ICPMS: List-1 (26)	6 Months	HNO3 to pH <2
B2M9W5	Y	W			1x500-mL G/P	6010_METALS_ICP: List-3 (18)	6 Months	HNO3 to pH <2
B2M9W5	Y	W			1x500-mL G	200.8_HG - ICPMS	28 Days	HNO3 to pH <2
B2M9W9	N	W			1x1-L aGs*	9020_TOX: TOX (1)	28 Days	H2SO4 to pH <2/Cool-4C
B2M9W9	N	W			1x250-mL aG	9060_TOC: TOC (1)	28 Days	HCl or H2SO4 to pH <2/Cool-4C
B2M9W8	N	W			1x1-L aGs*	9020_TOX: TOX (1)	28 Days	H2SO4 to pH <2/Cool-4C
B2M9W8	N	W			1x250-mL aG	9060_TOC: TOC (1)	28 Days	HCl or H2SO4 to pH <2/Cool-4C
B2M9W7	N	W			1x1-L aGs*	9020_TOX: TOX (1)	28 Days	H2SO4 to pH <2/Cool-4C
B2M9W7	N	W			1x250-mL aG	9060_TOC: TOC (1)	28 Days	HCl or H2SO4 to pH <2/Cool-4C
B2M9W6	N	W			1x500-mL G/P	200.8_METALS_ICPMS: List-1 (26)	6 Months	HNO3 to pH <2
B2M9W6	N	W			1x250-mL G/P	2320_ALKALINITY: Alkalinity (1)	14 Days	Cool-4C
B2M9W6	N	W			1x500-mL G/P	410.4_COD: COD (1)	28 Days	H2SO4 to pH <2/Cool-4C
B2M9W6	N	W			1x250-mL P	4500E_CN: Cyanide (1)	14 Days	NaOH to pH >=12
B2M9W6	N	W			1x500-mL G/P	6010_METALS_ICP: List-3 (18)	6 Months	HNO3 to pH <2

	Print	Sign	Date/Time	Received By	Print	Sign	Date/Time
Relinquished By: D.J. Woehle CHPRC			OCT 03 2012 1430		C Johnson		OCT 03 2012 1430
Relinquished By			Date/Time	Received By			Date/Time
Relinquished By			Date/Time	Received By			Date/Time
Relinquished By			Date/Time	Received By			Date/Time
FINAL SAMPLE DISPOSITION	Disposal Method (e.g., Return to customer, per lab procedure, used in process)				Disposed By		Date/Time

Matrix *

S = Soil	DS = Drum Solids		
SE = Sediment	DL = Drum Liquids		
SO = Solid	T = Tissue		
SL = Sludge	WI = Wipe		
W = Water	L = Liquid		
O = Oil	V = Vegetation		
A = Air	X = Other		

PRINTED ON 9/20/2012

A-6004-842 (REV 2)

Chain of Custody



CH2MHill Plateau Remediation Company

CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST

C.O.C. # W13-010-303

Page 2 of 2

Collector: D.J. Woehle, CHPRC
Contact/Requester: Karen Waters-Husted
Telephone No.: 376-4650
SAF No.: W13-010
Sampling Origin: Hanford Site
Purchase Order/Charge Code: 300071ES20
Project Title: RCRA, OCTOBER 2012
Logbook No.: HNF-N-506-48/68
Ice Chest No.: N/A
Shipped To (Lab): Waste Sampling & Characterization
Method of Shipment: GOVERNMENT VEHICLE
Bill of Lading/Air Bill No.: N/A
Protocol: RCRA
Priority: 31 Days
Offsite Property No.: N/A
PRIORITY

POSSIBLE SAMPLE HAZARDS/REMARKS
** ** Contains Radioactive Material at concentrations that are not regulated for transportation per 49 CRF but are not releasable per DOE Order 5400.5 (1990)(1993)

SPECIAL INSTRUCTIONS
FY12 and FY13 samples cannot be in the same SDG.
Site Wide Generator Knowledge Information Form applies.
The CACN for all analytical work at WSCF is 401647.

Hold Time
Total Activity Exemption: Yes ☑ No ☐

Sample No.	Filter	*	Date	Time	No/Type Container	Sample Analysis	Holding Time	Preservative
B2M9W6	N	W	10/3/12	1203	4x1-L aG	8082_PCB_GC: List-1 (7)	None	Cool-4C
B2M9W6	N	W			1x1-L aGs*	9020_TOX: TOX (1)	28 Days	H2SO4 to pH <2/Cool-4C
~~B2M9W6~~	~~N~~	~~W~~			~~1x250-mL O/P~~	~~0030_SULFIDE_Sulfur (1)~~	~~7 Days~~	~~ZnAc-NaOH to pH > 9/Cool-4C~~
B2M9W6	N	W			1x250-mL aG	9060_TOC: TOC (1)	28 Days	HCl or H2SO4 to pH <2/Cool-4C
B2M9W6	N	W			3x1-L aG	TPH-DieselKerosene Range - WTPH-D	14/40 Days	HCl to pH <2/Cool-4C
B2M9W6	N	W			4x40-mL aGs*	TPH-Gasoline Range - WTPH-G	14 Days	HCl to pH <2/Cool-4C
B2M9W6	N	W			1x500 mL G	200.8_HG - ICPMS	28 Days	HNO3 to pH <2
B2M9W6	N	W			3x40-mL aGs*	8260_VOA_GCMS_IX: COMMON; 8260_VOA_GCMS_IX: COMMON (Add-on)	14 Days	HCl or H2SO4 to pH <2/Cool-4C
B2M9W6	N	W			4x1-L aG	8270_SVOA_GCMS_IX: COMMON	7/40 Days	Cool-4C

Relinquished By: D.J. Woehle, CHPRC — Sign — Date/Time: OCT 0 3 2012 1203 — Received By — Print — Sign — Date/Time: OCT 0 3 2012 1203

Relinquished By — Date/Time — Received By — Date/Time

Relinquished By — Date/Time — Received By — Date/Time

Relinquished By — Date/Time — Received By — Date/Time

FINAL SAMPLE DISPOSITION: Disposal Method (e.g., Return to customer, per lab procedure, used in process) — Disposed By — Date/Time

Matrix *
S = Soil, DS = Drum Solids
SE = Sediment, DL = Drum Liquids
SO = Solid, T = Tissue
SL = Sludge, WI = Wipe
W = Water, L = Liquid
O = Oil, V = Vegetation
A = Air, X = Other

PRINTED ON 9/20/2012

A-6004-842 (REV 2)